




9

Worklist: 1491

<u>LAB CASE</u>	<u>ITEM</u>	<u>TASK ID</u>	<u>DESCRIPTION</u>	
C2017-0035	1	73662	AM 1 Urine Screening EMIT	
C2017-0036	1	73718	AM 1 Urine Screening EMIT	
C2017-0040	1	73854	AM 1 Urine Screening EMIT	
C2017-0055	1	74015	AM 1 Urine Screening EMIT	
C2017-0071	1	74255	AM 1 Urine Screening EMIT	
C2017-0083	1	74361	AM 1 Urine Screening EMIT	
C2017-0084	1	74364	AM 1 Urine Screening EMIT	
C2017-0091	1	74483	AM 1 Urine Screening EMIT	
C2017-0092	1	74487	AM 1 Urine Screening EMIT	
M2017-0024	2	74110	AM 1 Urine Screening EMIT	



AUTO MAINTENANCE

BLANK ROTOR RESULTS

Blank date/time: 1/17/2017 2:32:49 PM

Wavelength	Cuvette AV	Cuvette SD	Errors	Lamp AV	Lamp SD	Errors
340	0.4450	0.0187		1.9874	0.0006	
415	0.3423	0.0138		0.8527	0.0005	
505	0.2835	0.0126		0.2797	0.0005	
546	0.2670	0.0123		0.2024	0.0004	
570	0.2703	0.0120		0.2161	0.0004	
600	0.2625	0.0117		0.1164	0.0011	
660	0.2880	0.0118		0.0446	0.0006	
700	0.2811	0.0116		0.1555	0.0003	

CALIBRATION PLANNING

Test name	Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
+THC 50	DAT Cal	*								
+Amphet 500	DAT Cal	*								
+Benzodiazep300	DAT Cal	*								
+Cocaine 300	DAT Cal	*								
+Opiate 300	DAT Cal	*								

MAINTENANCE PLANNING

Maintenance procedure	0	1	2	3	4	5	6	7	>7 days valid
NEEDLE RINSE								*	
SYSTEM CLEAN									*

EXPIRING REAGENTS

Reagent name	0	1	2	3	4	5	6	7	>7 days valid

EXPIRING CONTROLS

Control name	0	1	2	3	4	5	6	7	>7 days valid
Level 0									*
Level 5									*
S2E									*
S2S									*

EXPIRING CALIBRATORS

Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
Acetaminophen									*
Salicylate Cal									*
sTOX Cal									*
sTCA Cal									*
ETOH Cal									*
DAT Cal									*
Ecstasy Cal 2									*
Ecstasy Cal 3									*
LSD Cal									*
qEcstasyCal 300									*
qEcstasyCal 500									*
qLSD Cal									*
cR Cal									*
sG Cal									*
pH Cal									*
cH Cal									*
nT Cal									*
oX Cal									*

AUTO MAINTENANCE

BLANK ROTOR RESULTS

Blank date/time: 1/17/2017 2:04:50 PM

Wavelength	Cuvette AV	Cuvette SD	Errors	Lamp AV	Lamp SD	Errors
340	369.1561	54.3313	UND SD.ERR	1.9879	0.0007	
415	0.3429	0.0136		0.8528	0.0007	
505	0.2839	0.0124		0.2804	0.0003	
546	0.2673	0.0122		0.2029	0.0004	
570	0.2706	0.0119		0.2160	0.0005	
600	0.2628	0.0116		0.1167	0.0012	
660	0.2882	0.0117		0.0446	0.0007	
700	0.2813	0.0115		0.1552	0.0006	

CALIBRATION PLANNING

Test name	Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
+THC 50	DAT Cal	*								
+Amphet 500	DAT Cal	*								
+Benzodiazep300	DAT Cal	*								
+Cocaine 300	DAT Cal	*								
+Opiate 300	DAT Cal	*								

MAINTENANCE PLANNING

Maintenance procedure	0	1	2	3	4	5	6	7	>7 days valid
NEEDLE RINSE								*	
SYSTEM CLEAN									*

EXPIRING REAGENTS

Reagent name	0	1	2	3	4	5	6	7	>7 days valid

EXPIRING CONTROLS

Control name	0	1	2	3	4	5	6	7	>7 days valid
Level 0									*
Level 5									*
S2E									*
S2S									*

EXPIRING CALIBRATORS

Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
Acetaminophen									*
Salicylate Cal									*
sTOX Cal									*
sTCA Cal									*
ETOH Cal									*
DAT Cal									*
Ecstasy Cal 2									*
Ecstasy Cal 3									*
LSD Cal									*
qEcstasyCal 300									*
qEcstasyCal 500									*
qLSD Cal									*
cR Cal									*
sG Cal									*
pH Cal									*
cH Cal									*
nT Cal									*
oX Cal									*

WORKLIST

Pos	Type	Rep	Sample data	Tests requested		
5	Sample	1	ID: brc3-68430 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
6	Sample	1	ID: brs3-67140 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
7	Sample	1	ID: neg-121616 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
8	Sample	1	ID: c2017-0036-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
9	Sample	1	ID: c2017-0040-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
10	Sample	1	ID: c2017-0055-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
11	Sample	1	ID: c2017-0071-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
12	Sample	1	ID: c2017-0083-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
13	Sample	1	ID: c2017-0084-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
14	Sample	1	ID: c2017-0091-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
15	Sample	1	ID: c2017-0092-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
16	Sample	1	ID: m2017-0024-2 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50

Measurement date: 1/17/2017 2:51:40 PM

Control name: Level 0

Batch number: 9a508ul-j3

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	negative	0.336 dAbs/m	0.051	dAbs/m	-
+Benzodiazep300	negative	0.287 dAbs/m	0.113	dAbs/m	-
+THC 50	negative	0.279 dAbs/m	0.052	dAbs/m	-
+Cocaine 300	negative	0.238 dAbs/m	0.108	dAbs/m	-
+Opiate 300	negative	0.190 dAbs/m	0.080	dAbs/m	-

Measurement date: 1/17/2017 2:53:55 PM

Control name: Level 5

Batch number: 9A609UL-h3

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.536 dAbs/m	0.148	dAbs/m	+
+Benzodiazep300	POSITIVE	0.512 dAbs/m	0.112	dAbs/m	+
+THC 50	POSITIVE	0.469 dAbs/m	0.138	dAbs/m	+
+Cocaine 300	POSITIVE	0.406 dAbs/m	0.060	dAbs/m	+
+Opiate 300	POSITIVE	0.430 dAbs/m	0.161	dAbs/m	+

Measurement date: 1/17/2017 2:55:43 PM

Control name: S2E

Batch number: 72520

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Benzodiazep300	POSITIVE	0.451 dAbs/m	0.052	dAbs/m	+
+THC 50	POSITIVE	0.412 dAbs/m	0.081	dAbs/m	+
+Cocaine 300	POSITIVE	0.377 dAbs/m	0.032	dAbs/m	+
+Opiate 300	POSITIVE	0.341 dAbs/m	0.072	dAbs/m	+

Measurement date: 1/17/2017 2:57:29 PM

Control name: S2S

Batch number: 60100

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.509 dAbs/m	0.121	dAbs/m	+



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen
Analytical Method 1.1

Laboratory Case No: brc3-68430
Date of Analysis: 1/17/2017
Time of Analysis: 3:11:53 PM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.399	0.388
Benzodiazepines (300 ng/mL)	POSITIVE 0.629	0.400
Cannabinoids (50 ng/mL)	negative 0.306	0.331
Cocaine Metabolite (300 ng/mL)	negative 0.280	0.345
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.443	0.269

VIVA-Jr Instrument ID: 8-1324





Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen
Analytical Method 1.1

Laboratory Case No: brs3-67140
Date of Analysis: 1/17/2017
Time of Analysis: 3:13:15 PM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.502	0.388
Benzodiazepines (300 ng/mL)	POSITIVE 0.640	0.400
Cannabinoids (50 ng/mL)	POSITIVE 0.437	0.331
Cocaine Metabolite (300 ng/mL)	POSITIVE 0.372	0.345
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.353	0.269

VIVA-Jr Instrument ID: 8-1324



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen
Analytical Method 1.1

Laboratory Case No: neg-121616
Date of Analysis: 1/17/2017
Time of Analysis: 3:15:30 PM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	negative 0.333	0.388
Benzodiazepines (300 ng/mL)	negative 0.283	0.400
Cannabinoids (50 ng/mL)	negative 0.283	0.331
Cocaine Metabolite (300 ng/mL)	negative 0.234	0.345
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	negative 0.191	0.269

VIVA-Jr Instrument ID: 8-1324