

Worklist: 1538

<u>LAB CASE</u>	<u>ITEM</u>	<u>TASK ID</u>	<u>DESCRIPTION</u>	
C2017-0226	1	76071	AM 1 Urine Screening EMIT	
C2017-0231	1	76149	AM 1 Urine Screening EMIT	
C2017-0235	1	76156	AM 1 Urine Screening EMIT	
C2017-0239	1	76231	AM 1 Urine Screening EMIT	
C2017-0252	1	76416	AM 1 Urine Screening EMIT	
C2017-0253	1	76419	AM 1 Urine Screening EMIT	
C2017-0260	1	76584	AM 1 Urine Screening EMIT	
M2017-0489	2	76235	AM 1 Urine Screening EMIT	
P2017-0232	2	76008	AM 1 Urine Screening EMIT	

AUTO MAINTENANCE**BLANK ROTOR RESULTS**

Blank date/time: 2/13/2017 10:07:23 AM

Wavelength	Cuvette AV	Cuvette SD	Errors	Lamp AV	Lamp SD	Errors
340	0.4479	0.0189		1.9883	0.0007	
415	0.3434	0.0139		0.8515	0.0007	
505	0.2842	0.0126		0.2804	0.0003	
546	0.2676	0.0124		0.2030	0.0004	
570	0.2708	0.0121		0.2163	0.0005	
600	0.2630	0.0118		0.1172	0.0012	
660	0.2884	0.0118		0.0453	0.0007	
700	0.2813	0.0116		0.1561	0.0007	

CALIBRATION PLANNING

Test name	Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
+THC 50	DAT Cal	*								
+Amphet 500	DAT Cal	*								
+Benzodiazep300	DAT Cal	*								
+Cocaine 300	DAT Cal	*								
+Opiate 300	DAT Cal	*								

MAINTENANCE PLANNING

Maintenance procedure	0	1	2	3	4	5	6	7	>7 days valid
NEEDLE RINSE								*	
SYSTEM CLEAN									*

EXPIRING REAGENTS

Reagent name	0	1	2	3	4	5	6	7	>7 days valid

EXPIRING CONTROLS

Control name	0	1	2	3	4	5	6	7	>7 days valid
Level 0									*
Level 5									*
S2E									*
S2S									*

EXPIRING CALIBRATORS

Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
Acetaminophen									*
Salicylate Cal									*
sTOX Cal									*
sTCA Cal									*
ETOH Cal									*
DAT Cal									*
Ecstasy Cal 2									*
Ecstasy Cal 3									*
LSD Cal									*
qEcstasyCal 300									*
qEcstasyCal 500									*
qLSD Cal									*
cR Cal									*
sG Cal									*
pH Cal									*
ch Cal									*
nT Cal									*
oX Cal									*

Measurement date: 2/13/2017 10:51:44 AM

Control name: Level 0

Batch number: 9a508ul-j3

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	negative	0.331 dAbs/m	0.062	dAbs/m	-
+Benzodiazep300	negative	0.285 dAbs/m	0.108	dAbs/m	-
+THC 50	negative	0.284 dAbs/m	0.042	dAbs/m	-
+Cocaine 300	negative	0.251 dAbs/m	0.115	dAbs/m	-
+Opiate 300	negative	0.189 dAbs/m	0.078	dAbs/m	-

Measurement date: 2/13/2017 10:53:59 AM

Control name: Level 5

Batch number: 9A609UL-J4

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.538 dAbs/m	0.144	dAbs/m	+
+Benzodiazep300	POSITIVE	0.481 dAbs/m	0.087	dAbs/m	+
+THC 50	POSITIVE	0.459 dAbs/m	0.133	dAbs/m	+
+Cocaine 300	POSITIVE	0.422 dAbs/m	0.056	dAbs/m	+
+Opiate 300	POSITIVE	0.417 dAbs/m	0.150	dAbs/m	+

Measurement date: 2/13/2017 10:55:47 AM

Control name: S2E

Batch number: 72520

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Benzodiazep300	POSITIVE	0.457 dAbs/m	0.063	dAbs/m	+
+THC 50	POSITIVE	0.448 dAbs/m	0.122	dAbs/m	+
+Cocaine 300	POSITIVE	0.423 dAbs/m	0.057	dAbs/m	+
+Opiate 300	POSITIVE	0.357 dAbs/m	0.089	dAbs/m	+

Measurement date: 2/13/2017 10:57:34 AM

Control name: S2S

Batch number: 60100

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.531 dAbs/m	0.138	dAbs/m	+

WORKLIST

Pos	Type	Rep	Sample data	Tests requested		
5	Sample	1	ID: neg 12517 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
6	Sample	1	ID: brc3-6843 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
7	Sample	1	ID: brs3-67140 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
8	Sample	1	ID: c2017-0226-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
9	Sample	1	ID: c2017-0231-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
10	Sample	1	ID: p2017-0232-2 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
11	Sample	1	ID: c2017-0235-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
12	Sample	1	ID: c2017-0239-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
13	Sample	1	ID: c2017-0252-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
14	Sample	1	ID: m2017-0489-2 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
15	Sample	1	ID: c2017-0253-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
16	Sample	1	ID: c2017-0260-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen
Analytical Method 1.1

Laboratory Case No: neg 12517
Date of Analysis: 2/13/2017
Time of Analysis: 12:58:43 PM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	negative 0.342	0.393
Benzodiazepines (300 ng/mL)	negative 0.294	0.394
Cannabinoids (50 ng/mL)	negative 0.287	0.326
Cocaine Metabolite (300 ng/mL)	negative 0.249	0.366
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	negative 0.189	0.267

VIVA-Jr Instrument ID: 8-1324



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen
Analytical Method 1.1

Laboratory Case No: brc3-6843
Date of Analysis: 2/13/2017
Time of Analysis: 1:00:57 PM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.414	0.393
Benzodiazepines (300 ng/mL)	POSITIVE 0.635	0.394
Cannabinoids (50 ng/mL)	negative 0.300	0.326
Cocaine Metabolite (300 ng/mL)	negative 0.308	0.366
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.448	0.267

VIVA-Jr Instrument ID: 8-1324



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen
Analytical Method 1.1

Laboratory Case No: brs3-67140
Date of Analysis: 2/13/2017
Time of Analysis: 1:10:24 PM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.505	0.393
Benzodiazepines (300 ng/mL)	POSITIVE 0.632	0.394
Cannabinoids (50 ng/mL)	POSITIVE 0.446	0.326
Cocaine Metabolite (300 ng/mL)	POSITIVE 0.395	0.366
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.353	0.267

VIVA-Jr Instrument ID: 8-1324