



Worklist: 1606

<u>LAB CASE</u>	<u>ITEM</u>	<u>TASK ID</u>	<u>DESCRIPTION</u>	
C2017-0443	1	78888	AM 1 Urine Screening EMIT	
C2017-0444	1	78892	AM 1 Urine Screening EMIT	
C2017-0445	1	78895	AM 1 Urine Screening EMIT	
C2017-0460	1	79117	AM 1 Urine Screening EMIT	
C2017-0463	1	79184	AM 1 Urine Screening EMIT	
C2017-0470	1	79230	AM 1 Urine Screening EMIT	
C2017-0472	1	79235	AM 1 Urine Screening EMIT	
C2017-0473	1	79238	AM 1 Urine Screening EMIT	
C2017-0486	1	79329	AM 1 Urine Screening EMIT	
C2017-0499	1	79461	AM 1 Urine Screening EMIT	
M2017-0949	3	78791	AM 1 Urine Screening EMIT	
M2017-1035	3	79045	AM 1 Urine Screening EMIT	



AUTO MAINTENANCE

BLANK ROTOR RESULTS

Blank date/time: 3/15/2017 7:02:09 AM

Wavelength	Cuvette AV	Cuvette SD	Errors	Lamp AV	Lamp SD	Errors
340	0.3872	0.0131		1.9346	0.0008	
415	0.3493	0.0065		0.7862	0.0008	
505	0.2570	0.0063		0.2179	0.0006	
546	0.2401	0.0062		0.1484	0.0005	
570	0.2449	0.0062		0.1592	0.0005	
600	0.2662	0.0061		0.0340	0.0015	
660	0.2652	0.0060		-0.0059	0.0007	
700	0.2903	0.0058		0.0628	0.0004	

CALIBRATION PLANNING

Test name	Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
+THC 50	DAT Cal	*								
+Amphet 500	DAT Cal	*								
+Benzodiazep300	DAT Cal	*								
+Cocaine 300	DAT Cal	*								
+Opiate 300	DAT Cal	*								

MAINTENANCE PLANNING

Maintenance procedure	0	1	2	3	4	5	6	7	>7 days valid
NEEDLE RINSE								*	
SYSTEM CLEAN									*

EXPIRING REAGENTS

Reagent name	0	1	2	3	4	5	6	7	>7 days valid

EXPIRING CONTROLS

Control name	0	1	2	3	4	5	6	7	>7 days valid
Level 0									*
Level 5									*
S2E									*
S2S									*

EXPIRING CALIBRATORS

Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
Acetaminophen									*
Salicylate Cal									*
sTOX Cal									*
sTCA Cal									*
ETOH Cal									*
DAT Cal									*
Ecstasy Cal 2									*
Ecstasy Cal 3									*
LSD Cal									*
qEcstasyCal 300									*
qEcstasyCal 500									*
qLSD Cal									*
cR Cal									*
sG Cal									*
pH Cal									*
cH Cal									*
nT Cal									*
oX Cal									*

BLANK/CALIBRATOR INFO

Test name	Reagent blank	Calibrator	
+Amphet 500		Accepted	Yes
		Value	0.3796 dAbs/m
		Meas. date	3/15/2017
		Type	One point - Cut-off
+Benzodiazep300		Accepted	Yes
		Value	0.4012 dAbs/m
		Meas. date	3/15/2017
		Type	One point - Cut-off
+THC 50		Accepted	Yes
		Value	0.3237 dAbs/m
		Meas. date	3/15/2017
		Type	One point - Cut-off
+Cocaine 300		Accepted	Yes
		Value	0.3630 dAbs/m
		Meas. date	3/15/2017
		Type	One point - Cut-off
+Opiate 300		Accepted	Yes
		Value	0.2688 dAbs/m
		Meas. date	3/15/2017
		Type	One point - Cut-off

Measurement date: 3/15/2017 7:55:37 AM

Control name: Level 0

Batch number: 9a508ul-j3

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	negative	0.326	dAbs/m	0.054	dAbs/m -
+Benzodiazep300	negative	0.284	dAbs/m	0.118	dAbs/m -
+THC 50	negative	0.284	dAbs/m	0.040	dAbs/m -
+Cocaine 300	negative	0.246	dAbs/m	0.117	dAbs/m -
+Opiate 300	negative	0.189	dAbs/m	0.080	dAbs/m -

Measurement date: 3/15/2017 7:57:52 AM

Control name: Level 5

Batch number: 9A609UL-J4

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.538	dAbs/m	0.158	dAbs/m +
+Benzodiazep300	POSITIVE	0.519	dAbs/m	0.117	dAbs/m +
+THC 50	POSITIVE	0.462	dAbs/m	0.138	dAbs/m +
+Cocaine 300	POSITIVE	0.440	dAbs/m	0.077	dAbs/m +
+Opiate 300	POSITIVE	0.433	dAbs/m	0.164	dAbs/m +

Measurement date: 3/15/2017 7:59:41 AM

Control name: S2E

Batch number: 72520

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Benzodiazep300	POSITIVE	0.457	dAbs/m	0.056	dAbs/m +
+THC 50	POSITIVE	0.422	dAbs/m	0.098	dAbs/m +
+Cocaine 300	POSITIVE	0.404	dAbs/m	0.041	dAbs/m +
+Opiate 300	POSITIVE	0.347	dAbs/m	0.078	dAbs/m +

Measurement date: 3/15/2017 8:01:26 AM

Control name: S2S

Batch number: 60100

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.511	dAbs/m	0.131	dAbs/m +

WORKLIST

Pos	Type	Rep	Sample data	Tests requested		
1	Sample	1	ID: brc3-68430 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
2	Sample	1	ID: brs3-67140 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
3	Sample	1	ID: neg-12517 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
4	Sample	1	ID: c2017-0443-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
5	Sample	1	ID: c2017-0444-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
6	Sample	1	ID: c2017-0445-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
7	Sample	1	ID: c2017-0460-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
8	Sample	1	ID: c2017-0463-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
9	Sample	1	ID: c2017-0470-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
10	Sample	1	ID: c2017-0472-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
11	Sample	1	ID: c2017-0473-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
12	Sample	1	ID: c2017-0486-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
13	Sample	1	ID: c2017-0499-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
14	Sample	1	ID: m2017-0949-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
15	Sample	1	ID: m2017-1035-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen
Analytical Method 1.1

Laboratory Case No: neg-12517
Date of Analysis: 3/15/2017
Time of Analysis: 10:45:21 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	negative 0.340	0.380
Benzodiazepines (300 ng/mL)	negative 0.285	0.401
Cannabinoids (50 ng/mL)	negative 0.278	0.324
Cocaine Metabolite (300 ng/mL)	negative 0.255	0.363
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	negative 0.188	0.269

VIVA-Jr Instrument ID: 8-1324



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen

Analytical Method 1.1

Laboratory Case No: brc3-68430
Date of Analysis: 3/15/2017
Time of Analysis: 10:40:50 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.407	0.380
Benzodiazepines (300 ng/mL)	POSITIVE 0.625	0.401
Cannabinoids (50 ng/mL)	negative 0.290	0.324
Cocaine Metabolite (300 ng/mL)	negative 0.300	0.363
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.451	0.269

VIVA-Jr Instrument ID: 8-1324



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen

Analytical Method 1.1

Laboratory Case No: brs3-67140
Date of Analysis: 3/15/2017
Time of Analysis: 10:43:05 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.503	0.380
Benzodiazepines (300 ng/mL)	POSITIVE 0.621	0.401
Cannabinoids (50 ng/mL)	POSITIVE 0.413	0.324
Cocaine Metabolite (300 ng/mL)	POSITIVE 0.384	0.363
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.348	0.269

VIVA-Jr Instrument ID: 8-1324