



Worklist: 1719

| <u>LAB_CASE</u> | <u>ITEM</u> | <u>TASK_ID</u> | <u>DESCRIPTION</u> | |
|-----------------|-------------|----------------|---------------------------|---|
| C2017-0806 | 1 | 83312 | AM 1 Urine Screening EMIT |  |
| C2017-0836 | 1 | 83704 | AM 1 Urine Screening EMIT |  |
| C2017-0837 | 1 | 83707 | AM 1 Urine Screening EMIT |  |
| C2017-0838 | 1 | 83710 | AM 1 Urine Screening EMIT |  |
| C2017-0852 | 1 | 83874 | AM 1 Urine Screening EMIT |  |
| C2017-0864 | 2 | 84192 | AM 1 Urine Screening EMIT |  |
| C2017-0885 | 2 | 84235 | AM 1 Urine Screening EMIT |  |
| M2017-1941 | 3 | 83745 | AM 1 Urine Screening EMIT |  |
| M2017-1942 | 3 | 83917 | AM 1 Urine Screening EMIT |  |
| M2017-2007 | 3 | 84023 | AM 1 Urine Screening EMIT |  |



AUTO MAINTENANCE**BLANK ROTOR RESULTS**

Blank date/time: 5/10/2017 7:37:31 AM

| Wavelength | Cuvette AV | Cuvette SD | Errors | Lamp AV | Lamp SD | Errors |
|------------|------------|------------|--------|---------|---------|--------|
| 340 | 0.3940 | 0.0134 | | 1.9406 | 0.0007 | |
| 415 | 0.3508 | 0.0060 | | 0.7863 | 0.0007 | |
| 505 | 0.2577 | 0.0057 | | 0.2194 | 0.0003 | |
| 546 | 0.2407 | 0.0055 | | 0.1488 | 0.0005 | |
| 570 | 0.2454 | 0.0055 | | 0.1595 | 0.0005 | |
| 600 | 0.2668 | 0.0054 | | 0.0367 | 0.0014 | |
| 660 | 0.2657 | 0.0053 | | -0.0053 | 0.0007 | |
| 700 | 0.2910 | 0.0052 | | 0.0619 | 0.0006 | |

CALIBRATION PLANNING

| Test name | Calibrator name | 0 | 1 | 2 | 3 | 4 | 5 | 6 | 7 | >7 days valid |
|-----------------|-----------------|---|---|---|---|---|---|---|---|---------------|
| +THC 50 | DAT Cal | * | | | | | | | | |
| +Amphet 500 | DAT Cal | * | | | | | | | | |
| +Benzodiazep300 | DAT Cal | * | | | | | | | | |
| +Cocaine 300 | DAT Cal | * | | | | | | | | |
| +Opiate 300 | DAT Cal | * | | | | | | | | |

MAINTENANCE PLANNING

| Maintenance procedure | 0 | 1 | 2 | 3 | 4 | 5 | 6 | 7 | >7 days valid |
|-----------------------|---|---|---|---|---|---|---|---|---------------|
| NEEDLE RINSE | | | | | | | | * | |
| SYSTEM CLEAN | | | | | | | | | * |

EXPIRING REAGENTS

| Reagent name | 0 | 1 | 2 | 3 | 4 | 5 | 6 | 7 | >7 days valid |
|--------------|---|---|---|---|---|---|---|---|---------------|
| | | | | | | | | | |

EXPIRING CONTROLS

| Control name | 0 | 1 | 2 | 3 | 4 | 5 | 6 | 7 | >7 days valid |
|--------------|---|---|---|---|---|---|---|---|---------------|
| Level 0 | | | | | | | | | * |
| Level 5 | | | | | | | | | * |
| S2E | | | | | | | | | * |
| S2S | | | | | | | | | * |

EXPIRING CALIBRATORS

| Calibrator name | 0 | 1 | 2 | 3 | 4 | 5 | 6 | 7 | >7 days valid |
|-----------------|---|---|---|---|---|---|---|---|---------------|
| Acetaminophen | | | | | | | | | * |
| Salicylate Cal | | | | | | | | | * |
| sTOX Cal | | | | | | | | | * |
| sTCA Cal | | | | | | | | | * |
| ETOH Cal | | | | | | | | | * |
| DAT Cal | | | | | | | | | * |
| Ecstasy Cal 2 | | | | | | | | | * |
| Ecstasy Cal 3 | | | | | | | | | * |
| LSD Cal | | | | | | | | | * |
| qEcstasyCal 300 | | | | | | | | | * |
| qEcstasyCal 500 | | | | | | | | | * |
| qLSD Cal | | | | | | | | | * |
| cR Cal | | | | | | | | | * |
| sG Cal | | | | | | | | | * |
| pH Cal | | | | | | | | | * |
| cH Cal | | | | | | | | | * |
| nT Cal | | | | | | | | | * |
| oX Cal | | | | | | | | | * |

Measurement date: 5/10/2017 8:02:37 AM

Control name: Level 0

Batch number: 9a508ul-j3

Sample type: Control READY

| Test name | Result | Units | Separation | Units | Flags |
|-----------------|----------|--------------|------------|--------|-------|
| +Amphet 500 | negative | 0.333 dAbs/m | 0.046 | dAbs/m | - |
| +Benzodiazep300 | negative | 0.282 dAbs/m | 0.120 | dAbs/m | - |
| +THC 50 | negative | 0.278 dAbs/m | 0.046 | dAbs/m | - |
| +Cocaine 300 | negative | 0.247 dAbs/m | 0.116 | dAbs/m | - |
| +Opiate 300 | negative | 0.186 dAbs/m | 0.082 | dAbs/m | - |

Measurement date: 5/10/2017 8:04:51 AM

Control name: Level 5

Batch number: 9A609UL-J4

Sample type: Control READY

| Test name | Result | Units | Separation | Units | Flags |
|-----------------|----------|--------------|------------|--------|-------|
| +Amphet 500 | POSITIVE | 0.536 dAbs/m | 0.156 | dAbs/m | + |
| +Benzodiazep300 | POSITIVE | 0.502 dAbs/m | 0.101 | dAbs/m | + |
| +THC 50 | POSITIVE | 0.462 dAbs/m | 0.138 | dAbs/m | + |
| +Cocaine 300 | POSITIVE | 0.423 dAbs/m | 0.060 | dAbs/m | + |
| +Opiate 300 | POSITIVE | 0.418 dAbs/m | 0.149 | dAbs/m | + |

Measurement date: 5/10/2017 8:06:40 AM

Control name: S2E

Batch number: 72520

Sample type: Control READY

| Test name | Result | Units | Separation | Units | Flags |
|-----------------|----------|--------------|------------|--------|-------|
| +Benzodiazep300 | POSITIVE | 0.453 dAbs/m | 0.052 | dAbs/m | + |
| +THC 50 | POSITIVE | 0.387 dAbs/m | 0.063 | dAbs/m | + |
| +Cocaine 300 | POSITIVE | 0.395 dAbs/m | 0.032 | dAbs/m | + |
| +Opiate 300 | POSITIVE | 0.341 dAbs/m | 0.072 | dAbs/m | + |

Measurement date: 5/10/2017 8:08:26 AM

Control name: S2S

Batch number: 60100

Sample type: Control READY

| Test name | Result | Units | Separation | Units | Flags |
|-------------|----------|--------------|------------|--------|-------|
| +Amphet 500 | POSITIVE | 0.472 dAbs/m | 0.093 | dAbs/m | + |

WORKLIST

| Pos | Type | Rep | Sample data | Tests requested |
|-----|--------|-----|---------------------------|--|
| 5 | Sample | 1 | ID: brc3-68440 Name: | 1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50 |
| 6 | Sample | 1 | ID: brs3-67140 Name: | 1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50 |
| 7 | Sample | 1 | ID: neg 42617 Name: | 1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50 |
| 8 | Sample | 1 | ID: c2017-0806-1 Name: | 1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50 |
| 9 | Sample | 1 | ID: c2017-0836-1 Name: | 1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50 |
| 10 | Sample | 1 | ID: c2017-0837-1 Name: | 1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50 |
| 11 | Sample | 1 | ID: c2017-0852-1 Name: | 1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50 |
| 12 | Sample | 1 | ID: c2017-0864-2 Name: | 1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50 |
| 13 | Sample | 1 | ID: c2017-0838-1 Name: | 1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50 |
| 14 | Sample | 1 | ID: m2017-2007-3 Name: | 1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50 |
| 15 | Sample | 1 | ID: m2017-1942-3 Name: | 1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50 |
| 16 | Sample | 1 | ID: c2017-0885-2 Name: | 1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50 |
| 17 | Sample | 1 | ID: m2017-1941-3 Name: | 1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50 |



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen
Analytical Method 1.1

Laboratory Case No: neg 42617
Date of Analysis: 5/10/2017
Time of Analysis: 8:30:57 AM

| Assay | Results/Interpretation | Cutoff Value (absorbance) |
|--------------------------------|------------------------|---------------------------|
| Amphetamines (500 ng/mL) | negative 0.327 | 0.380 |
| Benzodiazepines (300 ng/mL) | negative 0.281 | 0.401 |
| Cannabinoids (50 ng/mL) | negative 0.279 | 0.324 |
| Cocaine Metabolite (300 ng/mL) | negative 0.248 | 0.363 |
| Methadone (300 ng/mL) | | |
| Opiates (300 ng/mL) | negative 0.186 | 0.269 |

VIVA-Jr Instrument ID: 8-1324



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen
Analytical Method 1.1

Laboratory Case No: brs3-67140
Date of Analysis: 5/10/2017
Time of Analysis: 8:28:43 AM

| Assay | Results/Interpretation | Cutoff Value (absorbance) |
|--------------------------------|------------------------|---------------------------|
| Amphetamines (500 ng/mL) | POSITIVE 0.515 | 0.380 |
| Benzodiazepines (300 ng/mL) | POSITIVE 0.624 | 0.401 |
| Cannabinoids (50 ng/mL) | POSITIVE 0.389 | 0.324 |
| Cocaine Metabolite (300 ng/mL) | POSITIVE 0.389 | 0.363 |
| Methadone (300 ng/mL) | | |
| Opiates (300 ng/mL) | POSITIVE 0.361 | 0.269 |

VIVA-Jr Instrument ID: 8-1324



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen
Analytical Method 1.1

Laboratory Case No: brc3-68440
Date of Analysis: 5/10/2017
Time of Analysis: 8:26:27 AM

| Assay | Results/Interpretation | Cutoff Value (absorbance) |
|--------------------------------|------------------------|---------------------------|
| Amphetamines (500 ng/mL) | POSITIVE 0.408 | 0.380 |
| Benzodiazepines (300 ng/mL) | POSITIVE 0.623 | 0.401 |
| Cannabinoids (50 ng/mL) | negative 0.298 | 0.324 |
| Cocaine Metabolite (300 ng/mL) | negative 0.289 | 0.363 |
| Methadone (300 ng/mL) | | |
| Opiates (300 ng/mL) | POSITIVE 0.448 | 0.269 |

VIVA-Jr Instrument ID: 8-1324