

**Worklist: 1920**

<u>LAB_CASE</u>	<u>ITEM</u>	<u>TASK_ID</u>	<u>DESCRIPTION</u>	
C2017-1930	1	95686	AM 1 Urine Screening EMIT	
C2017-1943	1	96103	AM 1 Urine Screening EMIT	
C2017-1951	2	96030	AM 1 Urine Screening EMIT	
C2017-1951	3	96033	AM 1 Urine Screening EMIT	
M2017-4061	3	96106	AM 1 Urine Screening EMIT	
M2017-4093	4	96107	AM 1 Urine Screening EMIT	
M2017-4306	2	96108	AM 1 Urine Screening EMIT	
M2017-4351	3	96109	AM 1 Urine Screening EMIT	
M2017-4414	5	95678	AM 1 Urine Screening EMIT	
P2017-2229	1	96111	AM 1 Urine Screening EMIT	
P2017-2232	1	96112	AM 1 Urine Screening EMIT	
P2017-2267	1	96113	AM 1 Urine Screening EMIT	

# AUTO MAINTENANCE

## BLANK ROTOR RESULTS

Blank date/time: 9/27/2017 7:49:25 AM

Wavelength	Cuvette AV	Cuvette SD	Errors	Lamp AV	Lamp SD	Errors
340	0.3730	0.0140		1.9737	0.0007	
415	0.3226	0.0053		0.7998	0.0006	
505	0.2286	0.0051		0.2352	0.0004	
546	0.2120	0.0050		0.1668	0.0005	
570	0.2188	0.0050		0.1702	0.0005	
600	0.2409	0.0048		0.0486	0.0015	
660	0.2382	0.0048		0.0105	0.0007	
700	0.2645	0.0047		0.0756	0.0006	

## CALIBRATION PLANNING

Test name	Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
+THC 50	DAT Cal	*								
+Amphet 500	DAT Cal	*								
+Benzodiazep300	DAT Cal	*								
+Cocaine 300	DAT Cal	*								
+Opiate 300	DAT Cal	*								

## MAINTENANCE PLANNING

Maintenance procedure	0	1	2	3	4	5	6	7	>7 days valid
NEEDLE RINSE								*	
SYSTEM CLEAN									*

## EXPIRING REAGENTS

Reagent name	0	1	2	3	4	5	6	7	>7 days valid

## EXPIRING CONTROLS

Control name	0	1	2	3	4	5	6	7	>7 days valid
Level 0									*
Level 5									*
S2E									*
S2S									*

## EXPIRING CALIBRATORS

Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
Acetaminophen									*
Salicylate Cal									*
sTOX Cal									*
sTCA Cal									*
ETOH Cal									*
DAT Cal									*
Ecstasy Cal 2									*
Ecstasy Cal 3									*
LSD Cal									*
qEcstasyCal 300									*
qEcstasyCal 500									*
qLSD Cal									*
cR Cal									*
sG Cal									*
pH Cal									*
cH Cal									*
nT Cal									*
oX Cal									*

Measurement date: 9/27/2017 8:10:10 AM

Control name: Level 0

Batch number: 9a508ul-k1

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	negative	0.340	dAbs/m	0.070	-
+Benzodiazep300	negative	0.284	dAbs/m	0.110	-
+THC 50	negative	0.286	dAbs/m	0.044	-
+Cocaine 300	negative	0.255	dAbs/m	0.118	-
+Opiate 300	negative	0.189	dAbs/m	0.080	-

Measurement date: 9/27/2017 8:12:26 AM

Control name: Level 5

Batch number: 9A609UL-k1

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.530	dAbs/m	0.121	+
+Benzodiazep300	POSITIVE	0.489	dAbs/m	0.094	+
+THC 50	POSITIVE	0.476	dAbs/m	0.146	+
+Cocaine 300	POSITIVE	0.436	dAbs/m	0.063	+
+Opiate 300	POSITIVE	0.423	dAbs/m	0.154	+

Measurement date: 9/27/2017 8:14:14 AM

Control name: S2E

Batch number: 72520

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Benzodiazep300	POSITIVE	0.464	dAbs/m	0.069	+
+THC 50	POSITIVE	0.423	dAbs/m	0.093	+
+Cocaine 300	POSITIVE	0.425	dAbs/m	0.052	+
+Opiate 300	POSITIVE	0.355	dAbs/m	0.086	+

Measurement date: 9/27/2017 8:16:00 AM

Control name: S2S

Batch number: 60130

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.500	dAbs/m	0.091	+

**WORKLIST**

Pos	Type	Rep	Sample data	Tests requested
5	Sample	1	ID: brc3-68440 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
6	Sample	1	ID: brs2e-72520 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
7	Sample	1	ID: c2017-1930-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
8	Sample	1	ID: c2017-1943-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
9	Sample	1	ID: c2017-1951-2 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
10	Sample	1	ID: c2017-1951-3 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
11	Sample	1	ID: m2017-4061-3 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
12	Sample	1	ID: m2017-4093-4 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
13	Sample	1	ID: m2017-4306-2 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
14	Sample	1	ID: m2017-4351-3 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
15	Sample	1	ID: m2017-4414-5 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
16	Sample	1	ID: p2017-2229-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
17	Sample	1	ID: p2017-2232-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
18	Sample	1	ID: p2017-2367-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50



## Idaho State Police Forensic Services

**EMIT Urine Drugs-of-Abuse Screen**  
Analytical Method 1.1

Laboratory Case No: brc3-68440  
Date of Analysis: 9/27/2017  
Time of Analysis: 8:35:21 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.417	0.409
Benzodiazepines (300 ng/mL)	POSITIVE 0.619	0.395
Cannabinoids (50 ng/mL)	negative 0.319	0.330
Cocaine Metabolite (300 ng/mL)	negative 0.303	0.373
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.445	0.269

VIVA-Jr Instrument ID: 8-1324



## Idaho State Police Forensic Services

**EMIT Urine Drugs-of-Abuse Screen**  
Analytical Method 1.1

Laboratory Case No: brs2e-72520  
Date of Analysis: 9/27/2017  
Time of Analysis: 8:36:43 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.517	0.409
Benzodiazepines (300 ng/mL)	POSITIVE 0.406	0.395
Cannabinoids (50 ng/mL)	POSITIVE 0.350	0.330
Cocaine Metabolite (300 ng/mL)	POSITIVE 0.386	0.373
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.294	0.269

VIVA-Jr Instrument ID: 8-1324



## Idaho State Police Forensic Services

**EMIT Urine Drugs-of-Abuse Screen**  
Analytical Method 1.1

Laboratory Case No: neg-91917  
Date of Analysis: 9/27/2017  
Time of Analysis: 9:32:31 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	negative 0.339	0.409
Benzodiazepines (300 ng/mL)	negative 0.290	0.395
Cannabinoids (50 ng/mL)	negative 0.288	0.330
Cocaine Metabolite (300 ng/mL)	negative 0.258	0.373
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	negative 0.197	0.269

VIVA-Jr Instrument ID: 8-1324