
























*B. Wylee*

6/19/2017

**Worklist: 1771**

<u>LAB_CASE</u>	<u>ITEM</u>	<u>TASK_ID</u>	<u>DESCRIPTION</u>	
C2017-1025	1	85778	AM 1 Urine Screening EMIT	
C2017-1086	3	86651	AM 1 Urine Screening EMIT	
C2017-1097	1	86359	AM 1 Urine Screening EMIT	
C2017-1103	3	86399	AM 1 Urine Screening EMIT	
C2017-1104	1	86402	AM 1 Urine Screening EMIT	
C2017-1112	1	86497	AM 1 Urine Screening EMIT	
C2017-1126	1	86672	AM 1 Urine Screening EMIT	
C2017-1128	1	86680	AM 1 Urine Screening EMIT	
C2017-1129	1	86683	AM 1 Urine Screening EMIT	
C2017-1135	1	86704	AM 1 Urine Screening EMIT	
C2017-1157	1	86958	AM 1 Urine Screening EMIT	
C2017-1158	1	86961	AM 1 Urine Screening EMIT	
C2017-1172	1	86997	AM 1 Urine Screening EMIT	
C2017-1173	1	87002	AM 1 Urine Screening EMIT	
C2017-1182	1	87044	AM 1 Urine Screening EMIT	
C2017-1191	1	87108	AM 1 Urine Screening EMIT	
C2017-1194	1	87118	AM 1 Urine Screening EMIT	
C2017-1196	1	87162	AM 1 Urine Screening EMIT	
C2017-1215	1	87320	AM 1 Urine Screening EMIT	
C2017-1224	1	87414	AM 1 Urine Screening EMIT	
M2017-2147	2	87365	AM 1 Urine Screening EMIT	
M2017-2290	1	87366	AM 1 Urine Screening EMIT	
M2017-2702	5	87316	AM 1 Urine Screening EMIT	

*[Handwritten signature]*

**Worklist: 1771**

<u>LAB CASE</u>	<u>ITEM</u>	<u>TASK ID</u>	<u>DESCRIPTION</u>
P2017-1303	2	87323	AM 1 Urine Screening EMIT
P2017-1364	3	87327	AM 1 Urine Screening EMIT



A handwritten signature or mark, consisting of a stylized, overlapping loop and a trailing line, is located in the bottom right corner of the page.

**AUTO MAINTENANCE****BLANK ROTOR RESULTS**

Blank date/time: 6/20/2017 4:15:35 AM

Wavelength	Cuvette AV	Cuvette SD	Errors	Lamp AV	Lamp SD	Errors
340	0.4012	0.0067		1.9448	0.0008	
415	0.3155	0.0056		0.7870	0.0006	
505	0.2580	0.0054		0.2203	0.0004	
546	0.2407	0.0052		0.1498	0.0004	
570	0.2454	0.0052		0.1607	0.0005	
600	0.2670	0.0051		0.0391	0.0016	
660	0.2661	0.0051		-0.0056	0.0006	
700	0.2859	0.0050		0.0626	0.0002	

**CALIBRATION PLANNING**

Test name	Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
+THC 50	DAT Cal	*								
+Amphet 500	DAT Cal	*								
+Benzodiazep300	DAT Cal	*								
+Cocaine 300	DAT Cal	*								
+Opiate 300	DAT Cal	*								

**MAINTENANCE PLANNING**

Maintenance procedure	0	1	2	3	4	5	6	7	>7 days valid
NEEDLE RINSE								*	
SYSTEM CLEAN									*

**EXPIRING REAGENTS**

Reagent name	0	1	2	3	4	5	6	7	>7 days valid

**EXPIRING CONTROLS**

Control name	0	1	2	3	4	5	6	7	>7 days valid
Level 0									*
Level 5									*
S2E									*
S2S									*

**EXPIRING CALIBRATORS**

Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
Acetaminophen									*
Salicylate Cal									*
sTOX Cal									*
sTCA Cal									*
ETOH Cal									*
DAT Cal									*
Ecstasy Cal 2									*
Ecstasy Cal 3									*
LSD Cal									*
qEcstasyCal 300									*
qEcstasyCal 500									*
qLSD Cal									*
cR Cal									*
sG Cal									*
pH Cal									*
ch Cal									*
nT Cal									*
oX Cal									*

Measurement date: 6/20/2017 7:29:31 AM  
 Control name: Level 0  
 Batch number: 9a508ul-j3  
 Sample type: Control

READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	negative	0.344 dAbs/m		0.036 dAbs/m	-
+Benzodiazep300	negative	0.280 dAbs/m		0.121 dAbs/m	-
+THC 50	negative	0.282 dAbs/m		0.042 dAbs/m	-
+Cocaine 300	negative	0.244 dAbs/m		0.119 dAbs/m	-
+Opiate 300	negative	0.190 dAbs/m		0.079 dAbs/m	-

Measurement date: 6/20/2017 7:31:46 AM  
 Control name: Level 5  
 Batch number: 9A609UL-J4  
 Sample type: Control

READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.532 dAbs/m		0.152 dAbs/m	+
+Benzodiazep300	POSITIVE	0.497 dAbs/m		0.095 dAbs/m	+
+THC 50	POSITIVE	0.469 dAbs/m		0.145 dAbs/m	+
+Cocaine 300	POSITIVE	0.409 dAbs/m		0.046 dAbs/m	+
+Opiate 300	POSITIVE	0.422 dAbs/m		0.153 dAbs/m	+

Measurement date: 6/20/2017 7:33:34 AM  
 Control name: S2E  
 Batch number: 72520  
 Sample type: Control

READY

Test name	Result	Units	Separation	Units	Flags
+Benzodiazep300	POSITIVE	0.436 dAbs/m		0.034 dAbs/m	+
+THC 50	POSITIVE	0.374 dAbs/m		0.050 dAbs/m	+
+Cocaine 300	POSITIVE	0.391 dAbs/m		0.028 dAbs/m	+
+Opiate 300	POSITIVE	0.340 dAbs/m		0.072 dAbs/m	+

Measurement date: 6/20/2017 7:42:33 AM  
 Control name: S2S  
 Batch number: 60130  
 Sample type: Control

READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.432 dAbs/m		0.053 dAbs/m	+



## Idaho State Police Forensic Services

**EMIT Urine Drugs-of-Abuse Screen**  
Analytical Method 1.1

Laboratory Case No: brc3 68440  
Date of Analysis: 6/20/2017  
Time of Analysis: 8:40:52 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.408	0.380
Benzodiazepines (300 ng/mL)	POSITIVE 0.631	0.401
Cannabinoids (50 ng/mL)	negative 0.299	0.324
Cocaine Metabolite (300 ng/mL)	negative 0.295	0.363
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.440	0.269

VIVA-Jr Instrument ID: 8-1324



## Idaho State Police Forensic Services

**EMIT Urine Drugs-of-Abuse Screen**  
Analytical Method 1.1

Laboratory Case No: brs3 67140  
Date of Analysis: 6/20/2017  
Time of Analysis: 8:43:06 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.508	0.380
Benzodiazepines (300 ng/mL)	POSITIVE 0.618	0.401
Cannabinoids (50 ng/mL)	POSITIVE 0.387	0.324
Cocaine Metabolite (300 ng/mL)	POSITIVE 0.377	0.363
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.353	0.269

VIVA-Jr Instrument ID: 8-1324



## Idaho State Police Forensic Services

**EMIT Urine Drugs-of-Abuse Screen**  
Analytical Method 1.1

Laboratory Case No: neg 42617  
Date of Analysis: 6/20/2017  
Time of Analysis: 8:52:34 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	negative 0.337	0.380
Benzodiazepines (300 ng/mL)	negative 0.280	0.401
Cannabinoids (50 ng/mL)	negative 0.284	0.324
Cocaine Metabolite (300 ng/mL)	negative 0.241	0.363
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	negative 0.194	0.269

VIVA-Jr Instrument ID: 8-1324

**Nord, Anne**

---

**From:** Shrum, Celena  
**Sent:** Tuesday, June 20, 2017 6:05 AM  
**To:** Nord, Anne  
**Subject:** RE: expired reagents

Approved.

*Celena D. Shrum*



*Idaho State Police  
District 5 Forensics  
Forensic Scientist  
(208)232-9474  
celena.shrum@isp.idaho.gov*

**CONFIDENTIALITY NOTICE:** This e-mail is intended only for the personal and confidential use of the individual(s) named as recipients (or the employee or agent responsible to deliver it to the intended recipient) and is covered by the Electronic Communications Privacy Act, 18 U.S.C. §§ 2510-2521. It may contain information that is privileged, confidential and/or protected from disclosure under applicable law including, but not limited to, the attorney client privilege and/or work product doctrine. If you are not the intended recipient of this transmission, please notify the sender immediately by telephone. Do not deliver, distribute or copy this transmission, disclose its contents or take any action in reliance on the information it contains.

**From:** Nord, Anne  
**Sent:** Monday, June 19, 2017 5:32 PM  
**To:** Shrum, Celena <celena.shrum@isp.idaho.gov>  
**Subject:** expired reagents

Good Afternoon,

The quality manual allows for an expired reagent to be used when approved by discipline lead and tested with a positive and negative control.

*15.1.3.4.1 Reagents of questionable reliability and expired reagents shall be discarded. However, an expired reagent may continue to be used if tested with a positive and negative control each time it is used, and the appropriate discipline lead has approved the use of the expired reagent in writing before the release of results.*

I am requesting to use Amphetamine Assay lot J5 for toxicology AM#1 it expired 6/17/17. A positive and negative control will be run prior to the sample run, and an additional positive and negative control will be run within the sample run.

A handwritten signature in black ink, appearing to be "AS".