


























9/11/2017

**Worklist: 1887**

<u>LAB CASE</u>	<u>ITEM</u>	<u>TASK ID</u>	<u>DESCRIPTION</u>	
M2017-3448	2	94577	AM 1 Urine Screening EMIT	
M2017-3532	2	94578	AM 1 Urine Screening EMIT	
M2017-3538	3	94579	AM 1 Urine Screening EMIT	
M2017-3791	3	94580	AM 1 Urine Screening EMIT	
P2017-1746	1	94581	AM 1 Urine Screening EMIT	
P2017-1747	1	94589	AM 1 Urine Screening EMIT	
P2017-1755	1	94590	AM 1 Urine Screening EMIT	
P2017-1760	1	94591	AM 1 Urine Screening EMIT	
P2017-1788	1	94592	AM 1 Urine Screening EMIT	
P2017-1789	1	94593	AM 1 Urine Screening EMIT	
P2017-1804	1	94594	AM 1 Urine Screening EMIT	
P2017-1820	4	94595	AM 1 Urine Screening EMIT	
P2017-1825	1	94596	AM 1 Urine Screening EMIT	
P2017-1830	2	94597	AM 1 Urine Screening EMIT	
P2017-1832	1	94608	AM 1 Urine Screening EMIT	
P2017-1833	1	94598	AM 1 Urine Screening EMIT	
P2017-1834	1	94599	AM 1 Urine Screening EMIT	
P2017-1835	1	94600	AM 1 Urine Screening EMIT	
P2017-1845	1	94609	AM 1 Urine Screening EMIT	
P2017-1865	1	94601	AM 1 Urine Screening EMIT	
P2017-1878	1	94602	AM 1 Urine Screening EMIT	
P2017-1880	1	94603	AM 1 Urine Screening EMIT	
P2017-1881	1	94610	AM 1 Urine Screening EMIT	

**Worklist: 1887**

<u>LAB CASE</u>	<u>ITEM</u>	<u>TASK ID</u>	<u>DESCRIPTION</u>	
P2017-1897	1	94611	AM 1 Urine Screening EMIT	
P2017-1902	1	94604	AM 1 Urine Screening EMIT	
P2017-1903	1	94605	AM 1 Urine Screening EMIT	
P2017-1905	1	94606	AM 1 Urine Screening EMIT	
P2017-1949	1	94607	AM 1 Urine Screening EMIT	
P2017-1949	2	94588	AM 1 Urine Screening EMIT	
P2017-1951	1	94587	AM 1 Urine Screening EMIT	
P2017-1952	1	94586	AM 1 Urine Screening EMIT	
P2017-1954	1	94585	AM 1 Urine Screening EMIT	
P2017-1959	1	94584	AM 1 Urine Screening EMIT	
P2017-1964	1	94583	AM 1 Urine Screening EMIT	
P2017-2009	1	94582	AM 1 Urine Screening EMIT	



**AUTO MAINTENANCE****BLANK ROTOR RESULTS**

Blank date/time: 9/8/2017 6:41:10 PM

Wavelength	Cuvette AV	Cuvette SD	Errors	Lamp AV	Lamp SD	Errors
340	0.3992	0.0166		1.9501	0.0007	
415	0.3531	0.0062		0.7854	0.0006	
505	0.2598	0.0057		0.2212	0.0003	
546	0.2426	0.0056		0.1501	0.0005	
570	0.2474	0.0056		0.1612	0.0005	
600	0.2687	0.0055		0.0352	0.0015	
660	0.2676	0.0054		-0.0043	0.0007	
700	0.2928	0.0052		0.0624	0.0006	

**CALIBRATION PLANNING**

Test name	Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
+THC 50	DAT Cal	*								
+Amphet 500	DAT Cal	*								
+Benzodiazep300	DAT Cal	*								
+Cocaine 300	DAT Cal	*								
+Opiate 300	DAT Cal	*								

**MAINTENANCE PLANNING**

Maintenance procedure	0	1	2	3	4	5	6	7	>7 days valid
NEEDLE RINSE								*	
SYSTEM CLEAN	*								

**EXPIRING REAGENTS**

Reagent name	0	1	2	3	4	5	6	7	>7 days valid

**EXPIRING CONTROLS**

Control name	0	1	2	3	4	5	6	7	>7 days valid
Level 0									*
Level 5									*
S2E									*
S2S									*

**EXPIRING CALIBRATORS**

Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
Acetaminophen									*
Salicylate Cal									*
sTOX Cal									*
sTCA Cal									*
ETOH Cal									*
DAT Cal									*
Ecstasy Cal 2									*
Ecstasy Cal 3									*
LSD Cal									*
qEcstasyCal 300									*
qEcstasyCal 500									*
qLSD Cal									*
cR Cal									*
sG Cal									*
pH Cal									*
cH Cal									*
nT Cal									*
oX Cal									*

**BLANK/CALIBRATOR INFO**

Test name	Reagent blank	Calibrator	
+Amphet 500		Accepted	Yes
		Value	0.4598 dAbs/m
		Meas. date	9/11/2017
		Type	One point - Cut-off
+Benzodiazep300		Accepted	Yes
		Value	0.4285 dAbs/m
		Meas. date	9/11/2017
		Type	One point - Cut-off
+THC 50		Accepted	Yes
		Value	0.3585 dAbs/m
		Meas. date	9/11/2017
		Type	One point - Cut-off
+Cocaine 300		Accepted	Yes
		Value	0.3764 dAbs/m
		Meas. date	9/11/2017
		Type	One point - Cut-off
+Opiate 300		Accepted	Yes
		Value	0.3180 dAbs/m
		Meas. date	9/11/2017
		Type	One point - Cut-off

Measurement date: 9/11/2017 8:47:59 AM

Control name: Level 0

Batch number: 9a508ul-k1

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	negative	0.336 dAbs/m	0.124	dAbs/m	-
+Benzodiazep300	negative	0.286 dAbs/m	0.142	dAbs/m	-
+THC 50	negative	0.295 dAbs/m	0.063	dAbs/m	-
+Cocaine 300	negative	0.246 dAbs/m	0.130	dAbs/m	-
+Opiate 300	negative	0.194 dAbs/m	0.124	dAbs/m	-

Measurement date: 9/11/2017 8:50:14 AM

Control name: Level 5

Batch number: 9A609UL-k1

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.525 dAbs/m	0.065	dAbs/m	+
+Benzodiazep300	POSITIVE	0.492 dAbs/m	0.064	dAbs/m	+
+THC 50	POSITIVE	0.476 dAbs/m	0.117	dAbs/m	+
+Cocaine 300	POSITIVE	0.414 dAbs/m	0.037	dAbs/m	+
+Opiate 300	POSITIVE	0.419 dAbs/m	0.101	dAbs/m	+

Measurement date: 9/11/2017 8:52:02 AM

Control name: S2E

Batch number: 72520

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Benzodiazep300	POSITIVE	0.448 dAbs/m	0.020	dAbs/m	+
+THC 50	POSITIVE	0.411 dAbs/m	0.052	dAbs/m	+
+Cocaine 300	POSITIVE	0.382 dAbs/m	0.006	dAbs/m	+
+Opiate 300	POSITIVE	0.337 dAbs/m	0.019	dAbs/m	+

Measurement date: 9/11/2017 8:53:49 AM

Control name: S2S

Batch number: 60130

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.485 dAbs/m	0.025	dAbs/m	+

BW

## WORKLIST

Pos	Type	Rep	Sample data	Tests requested
1	Sample	1	ID: brc3-68440 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
2	Sample	1	ID: brs2e-72520 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
3	Sample	1	ID: neg-9517 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
4	Sample	1	ID: p2017-1746-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
5	Sample	1	ID: p2017-1747-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
6	Sample	1	ID: p2017-1755-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
7	Sample	1	ID: p2017-1760-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
8	Sample	1	ID: p2017-1788-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
9	Sample	1	ID: p2017-1789-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
10	Sample	1	ID: p2017-1804-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
11	Sample	1	ID: p2017-1825-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
12	Sample	1	ID: p2017-1830-2 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
13	Sample	1	ID: p2017-1832-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
14	Sample	1	ID: p2017-1833-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
15	Sample	1	ID: p2017-1834-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
16	Sample	1	ID: p2017-1835-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
17	Sample	1	ID: p2017-1845-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
18	Sample	1	ID: p2017-1865-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
19	Sample	1	ID: p2017-1878-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
20	Sample	1	ID: p2017-1880-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
21	Sample	1	ID: p2017-1881-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
22	Sample	1	ID: m2017-3448-2 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
23	Sample	1	ID: m2017-3532-2 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
24	Sample	1	ID: m2017-3538-3 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
25	Sample	1	ID: m2017-3791-3 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50



## Idaho State Police Forensic Services

**EMIT Urine Drugs-of-Abuse Screen**  
Analytical Method 1.1

Laboratory Case No: brc3-68440  
Date of Analysis: 9/11/2017  
Time of Analysis: 9:32:14 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.495	0.460
Benzodiazepines (300 ng/mL)	POSITIVE 0.630	0.429
Cannabinoids (50 ng/mL)	negative 0.298	0.359
Cocaine Metabolite (300 ng/mL)	negative 0.317	0.376
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.445	0.318

VIVA-Jr Instrument ID: 8-1324



## Idaho State Police Forensic Services

**EMIT Urine Drugs-of-Abuse Screen**  
Analytical Method 1.1

Laboratory Case No: brs2e-72520  
Date of Analysis: 9/11/2017  
Time of Analysis: 9:34:29 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.538	0.460
Benzodiazepines (300 ng/mL)	POSITIVE 0.456	0.429
Cannabinoids (50 ng/mL)	POSITIVE 0.395	0.359
Cocaine Metabolite (300 ng/mL)	POSITIVE 0.386	0.376
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.343	0.318

VIVA-Jr Instrument ID: 8-1324





## Idaho State Police Forensic Services

**EMIT Urine Drugs-of-Abuse Screen**  
Analytical Method 1.1

Laboratory Case No: neg-9517  
Date of Analysis: 9/11/2017  
Time of Analysis: 9:43:56 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	negative 0.335	0.460
Benzodiazepines (300 ng/mL)	negative 0.286	0.429
Cannabinoids (50 ng/mL)	negative 0.290	0.359
Cocaine Metabolite (300 ng/mL)	negative 0.248	0.376
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	negative 0.193	0.318

VIVA-Jr Instrument ID: 8-1324