





**Worklist: 1592**

<u>LAB CASE</u>	<u>ITEM</u>	<u>TASK ID</u>	<u>DESCRIPTION</u>
C2017-0345	1	77652	AM 1 Urine Screening EMIT
C2017-0365	2	77928	AM 1 Urine Screening EMIT
C2017-0366	1	77931	AM 1 Urine Screening EMIT
C2017-0372	1	77940	AM 1 Urine Screening EMIT
C2017-0373	1	77961	AM 1 Urine Screening EMIT
C2017-0403	1	78501	AM 1 Urine Screening EMIT
C2017-0408	1	78540	AM 1 Urine Screening EMIT
C2017-0410	2	78547	AM 1 Urine Screening EMIT
C2017-0428	1	78752	AM 1 Urine Screening EMIT
C2017-0429	1	78756	AM 1 Urine Screening EMIT
M2017-0920	3	78449	AM 1 Urine Screening EMIT
P2017-0404	2	77696	AM 1 Urine Screening EMIT



# AUTO MAINTENANCE

## BLANK ROTOR RESULTS

Blank date/time: 3/8/2017 8:41:13 AM

Wavelength	Cuvette AV	Cuvette SD	Errors	Lamp AV	Lamp SD	Errors
340	0.4475	0.0183		1.9884	0.0011	
415	0.3429	0.0136		0.8511	0.0006	
505	0.2839	0.0124		0.2797	0.0004	
546	0.2671	0.0122		0.2023	0.0004	
570	0.2702	0.0119		0.2160	0.0004	
600	0.2626	0.0116		0.1153	0.0010	
660	0.2881	0.0117		0.0441	0.0004	
700	0.2809	0.0115		0.1563	0.0004	

## CALIBRATION PLANNING

Test name	Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
+THC 50	DAT Cal	*								
+Amphet 500	DAT Cal	*								
+Benzodiazep300	DAT Cal	*								
+Cocaine 300	DAT Cal	*								
+Opiate 300	DAT Cal	*								

## MAINTENANCE PLANNING

Maintenance procedure	0	1	2	3	4	5	6	7	>7 days valid
NEEDLE RINSE								*	
SYSTEM CLEAN									*

## EXPIRING REAGENTS

Reagent name	0	1	2	3	4	5	6	7	>7 days valid

## EXPIRING CONTROLS

Control name	0	1	2	3	4	5	6	7	>7 days valid
Level 0									*
Level 5									*
S2E									*
S2S									*

## EXPIRING CALIBRATORS

Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
Acetaminophen									*
Salicylate Cal									*
sTOX Cal									*
sTCA Cal									*
ETOH Cal									*
DAT Cal									*
Ecstasy Cal 2									*
Ecstasy Cal 3									*
LSD Cal									*
qEcstasyCal 300									*
qEcstasyCal 500									*
qLSD Cal									*
cR Cal									*
sG Cal									*
pH Cal									*
cH Cal									*
nT Cal									*
oX Cal									*

*BW*

Measurement date: 3/8/2017 9:09:03 AM  
Control name: Level 0  
Batch number: 9a508ul-j3  
Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	negative	0.341 dAbs/m	0.052	dAbs/m	-
+Benzodiazep300	negative	0.286 dAbs/m	0.108	dAbs/m	-
+THC 50	negative	0.290 dAbs/m	0.036	dAbs/m	-
+Cocaine 300	negative	0.252 dAbs/m	0.115	dAbs/m	-
+Opiate 300	negative	0.191 dAbs/m	0.076	dAbs/m	-

Measurement date: 3/8/2017 9:11:17 AM  
Control name: Level 5  
Batch number: 9A609UL-J4  
Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.549 dAbs/m	0.156	dAbs/m	+
+Benzodiazep300	POSITIVE	0.515 dAbs/m	0.122	dAbs/m	+
+THC 50	POSITIVE	0.476 dAbs/m	0.150	dAbs/m	+
+Cocaine 300	POSITIVE	0.436 dAbs/m	0.070	dAbs/m	+
+Opiate 300	POSITIVE	0.433 dAbs/m	0.166	dAbs/m	+

Measurement date: 3/8/2017 9:13:05 AM

Control name: S2E

Batch number: 72520

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Benzodiazep300	POSITIVE	0.455 dAbs/m	0.061	dAbs/m	+
+THC 50	POSITIVE	0.408 dAbs/m	0.082	dAbs/m	+
+Cocaine 300	POSITIVE	0.404 dAbs/m	0.038	dAbs/m	+
+Opiate 300	POSITIVE	0.352 dAbs/m	0.085	dAbs/m	+

Measurement date: 3/8/2017 9:22:04 AM

Control name: S2S

Batch number: 60100

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.516 dAbs/m	0.123	dAbs/m	+

**WORKLIST**

Pos	Type	Rep	Sample data	Tests requested
5	Sample	1	ID: brc3-68430 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
6	Sample	1	ID: brs3-67140 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
7	Sample	1	ID: neg-12517 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
8	Sample	1	ID: c2017-0345-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
9	Sample	1	ID: c2017-0365-2 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
10	Sample	1	ID: c2017-0366-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
11	Sample	1	ID: c2017-0372-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
12	Sample	1	ID: c2017-0403-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
13	Sample	1	ID: c2017-0408-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
14	Sample	1	ID: c2017-0410-2 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
15	Sample	1	ID: m2017-0920-3 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
16	Sample	1	ID: p2017-0404-2 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50

**WORKLIST**

Pos	Type	Rep	Sample data	Tests requested		
17	Sample	1	ID: c2017-0428-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
18	Sample	1	ID: c2017-0429-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
19	Sample	1	ID: c2017-0373-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50



## Idaho State Police Forensic Services

**EMIT Urine Drugs-of-Abuse Screen**  
Analytical Method 1.1

Laboratory Case No: neg-12517  
Date of Analysis: 3/8/2017  
Time of Analysis: 9:30:11 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	negative 0.351	0.393
Benzodiazepines (300 ng/mL)	negative 0.290	0.394
Cannabinoids (50 ng/mL)	negative 0.288	0.326
Cocaine Metabolite (300 ng/mL)	negative 0.251	0.366
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	negative 0.194	0.267

VIVA-Jr Instrument ID: 8-1324

BW



## Idaho State Police Forensic Services

**EMIT Urine Drugs-of-Abuse Screen**  
Analytical Method 1.1

Laboratory Case No: brs3-67140  
Date of Analysis: 3/8/2017  
Time of Analysis: 9:27:29 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.512	0.393
Benzodiazepines (300 ng/mL)	POSITIVE 0.625	0.394
Cannabinoids (50 ng/mL)	POSITIVE 0.451	0.326
Cocaine Metabolite (300 ng/mL)	POSITIVE 0.397	0.366
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.363	0.267

VIVA-Jr Instrument ID: 8-1324

BW





## Idaho State Police Forensic Services

**EMIT Urine Drugs-of-Abuse Screen**  
Analytical Method 1.1

Laboratory Case No: brc3-68430  
Date of Analysis: 3/8/2017  
Time of Analysis: 9:25:15 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.415	0.393
Benzodiazepines (300 ng/mL)	POSITIVE 0.624	0.394
Cannabinoids (50 ng/mL)	negative 0.304	0.326
Cocaine Metabolite (300 ng/mL)	negative 0.302	0.366
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.451	0.267

VIVA-Jr Instrument ID: 8-1324

*BW*