

5.0 Balance Calibration and Intermediate Checks

5.1 BACKGROUND

Refer to manufacturer's manual for balance specific procedures.

5.2 SCOPE

The procedure is to ensure accuracy of measurement when an analytical or top-loading balance is used to prepare solutions or reference material for application to toxicology methods.

5.3 EQUIPMENT

- 5.3.1 ANSI/ASTM Type I, Class 1 or Class 2 laboratory weights
- 5.3.2 Analytical and/or Top-Loading, Direct-Reading Laboratory Balances

5.4 QUALITY ASSURANCE REQUIREMENTS

- 5.4.1 All balances will be calibrated and serviced yearly by a qualified vendor.
- 5.4.2 Weights used for intermediate checks of calibration will be calibrated yearly by a qualified vendor.
- An in-house intermediate check of balances in use will be performed on a monthly basis. Results are to be recorded in logbook. Balances not in-use need not be checked, however, the balance must then be checked prior to use.
- The weights used for the intermediate check should depend on the application of the balance. Three weights must be used to represent the weight range in question. For instance, if the balance is being used to prepare buffer solutions, then perhaps 100g, 500g and 1000g weights would characterize the weight range. If the balance is used to prepare mg/mL reference material then 1mg, 10mg and 100mg/mL weights may be appropriate.
- 5.4.5 The type of balance employed should be a consideration. For weights less than 100mg, an analytical balance should be used. For the preparation of a solution involving gram quantities, a top-

loading balance should be used.

Do not touch weights with bare hands. Weights should be transferred with forceps or gloves.

5. 5 GENERAL PROCEDURES

- 5.5.1 Inspect balance pan, clean if necessary
- 5.5.2 Inspect level bubble, level if necessary
- 5.5.3 Tare balance with weighing paper
- 5.5.4 Place weight on balance
- 5.5.5 If appropriate, add or subtract correction factor for weight as determined by yearly weight calibration certificates.
- 5.5.6 Record corrected weight on balance logsheet.
- 5.5.7 The acceptable range for the balance is based on whether an analytical or top-loading balance is used.
 - 5.5.7.1 Analytical Balance

Weight	Tolerance	Acceptable Range
	±1.0mg	99.0mg - 101.0g
(0.1g)	(0.001g)	
100g	±0.1g	99.9g - 100.1g

5.5.7.2 Top-loading Balance

Weight	Tolerance	Acceptable Range
10.00g	±0.02g	9.98g - 10.02g
100.00g	±0.2g	99.8g - 100.2g
500g	±0.5g	499.5g - 500.5g
1000g	±1.0g	999.0g - 1001.0g

5.8 The verification procedure should be repeated if the corrected value does not fall within the acceptable range. If value is still out of range, contact service vendor to set up a service call. A note must be placed on the balance to indicate that it is not in range and must not be used.

5. 6 REFERENCES

5.6.1 ASTM Method E-617-97, Standard Specification for Laboratory Weights and Precision Mass Standards.

Rev. 0 Issued: 1-20-2011 5.0 – Volatiles Balance Requirements – Rev 0.doc Issuing Authority: Quality Manager

Revision History

5.0 Balance Calibration and Intermediate Checks

	Revision #	Issue	Date	History	
	1	09-18-2001	Original Iss	ue	ices
	2	05-07-2007	requirement	ts	reformatting based on current
Q ⁴	Operity	01-20-2011	Initial version analytical m	on as 1.0, split freethods.	from toxicology discipline

topethy of Idaho state Police Forest Constitution of Idaho state Police Internet Int

Gamette, Matthew

From: Johnston, Jeremy

Friday, August 19, 2011 1:04 PM Sent:

Gamette, Matthew To:

Subject: rescind

AM 2.0, 3.0, 5.0 and 9.0 are no longer in use or have been incorporated into other AM's and can be archived as no longer in use.

AM 2.0 has been incorporated into AM 1.0

AM 9.0 has been incorporate into AM 8.0

AM 5.0 is redundant to the requirements of Tox or Drug AM's and referenced as such.

AM 3.0 is redundant with Tox AM's

Jeremy Johnston Alcohol Discipline Leader **Idaho State Police Forensics** Coeur d' Alene ID 208-209-8706

