

10.0 Uncertainty of Measurement for Volatiles Analysis

10.1 BACKGROUND

Any measurement, no matter how carefully obtained, should not be considered as the true value for the measurement. Whenever any quantitative measurement is performed, the value obtained is only an approximation of the true value. According to JCGM 200:2008, the International vocabulary of metrology – Basic and general concepts and associated terms (VIM), measurement uncertainty is defined as "A non-negative parameter associated with the result of a measurement/quantity value (number and measurement unit used together to express the magnitude of a quantity) that characterizes the dispersion of quantity values that could reasonably be attributed to the measurand (quantity intended to be measured)." ISO/IEC 17025:2005 clause 5.4.6.2 requires that we make a reasonable estimation of uncertainty that is based on knowledge of the performance of the method and on the measurement scope and shall make use of for example, previous experience and validation data.² Clause 5.4.6.2. NOTE 1 goes on to state that the degree of rigor needed in an estimation of uncertainty of measurement depends on factors such as the existence of narrow limits on which decisions on conformity to a specification is based.² Paragraph 5.10.3.1 states that when applicable, the test report should include a statement on the estimated uncertainty of measurement.² For our purposes, it is applicable due to the uncertainty affecting the application of the test results which are compliant to a specification limit. In the analysis of forensic specimens, we do not know the true value for the specimen; hence this information is not the error associated with the analysis. Rather, it is a range of values likely to be encountered during the measurement process. This information is crucial to the legal system because it impacts if and how an individual will be charged with an offense such as DUI.^{4,5}

10.2 SCOPE

This analytical method will be applied to analytical methods which report quantitative results. This approach to uncertainty uses the standard deviation of matrix matched controls and other known sources of uncertainty. A 99% confidence interval will be created by three standard deviations of data collected during the process. To properly represent the uncertainty, this data will be expressed as the Uncertainty Of Measurement on the analysis report. Authentication of ethanol containing blood controls is described in Volatiles Analysis Analytical Method 2.0

10.3 EQUIPMENT

Reference analytical methods listed under section 10.6.

10.4 REAGENTS

Reference analytical methods listed under section 10.6.

10.5 QUALITY ASSURANCE MATERIAL

Reference analytical methods listed under section 10.6.

10.6 REPORTING OF QUANTITATIVE ETHANOL RESULTS

10.6.1 Analytical Methods

1.0 Analysis of Volatiles by GC-HS

10.6.2 Determination of Confidence Interval

10.6.2.1 Blood control values obtained during the process are used to establish the UM based on the standard deviation of data as well as incorporating other known sources of uncertainty into the uncertainty budget.

Three standard deviations will be calculated for a 99% confidence interval.

The mean value as determined by the above analytical method will be reported along with a \pm UM.

10.70 REFERENCES AND RECOMMENDED READING

- 10.7.1 Huber, L., Validation and Qualification in Analytical Laboratories, pp. 146 150, Interpharm/CRC, 19910.
- 10.7.2 International Organization of Standardization (ISO) / International Electrochemical Commission (IEC), General requirements for the competence of testing and calibration laboratories, 2005. (ISO/IEC 17025:2005)
- 10.7.3 Joint Committee for Guides in Metrology (JCGM), *International Vocabulary of Basic and General Terms in Metrology (VIM)*, 2008. (JCGM 200: 2008)

- 10.7.4 Idaho Code §18-8004. Persons under the influence of alcohol, drugs or any other intoxicating substances.
- 10.7.5 Idaho Code §18-8004C. Excessive Alcohol Concentration -Penalties.
- 10.7.6 ISO/IEC 17025:2005: Section 5.4.6: Estimation of Uncertainty of ainty, Quality E

 aver, Col

 ainty, Quality E

 property of Idahochtrollocularity in the Id Measurement Workshop, Presented by J.P. Bono and E.A. Mishalanie, AAFS 61st Annual Meeting, Denver, Colorado, 20010.
 - Mason, F., Uncertain About Uncertainty, Quality Digest, Inside



Revision History

10.0 Uncertainty of Measurement for Volatiles Analysis

	Revision #	Issue Date	Revisions
	0	09-07-2009	Original issue Analytical Methods 4.1 and 4.2 addressed for quantitative ethanol results.
	0	1-20-2011	Initial version as 10.0, split from toxicology discipline analytical methods. Formerly Toxicology AM 5.1.13.
	1	4-23-2012	Changes made to reflect correct references to other AM's.
<i>Q</i> ⁴	Operty	4-15-2013	Changes made to sections 10.1, 10.2, 10.6.2.1, 10.6.2.2, 10.6.2.3.