











Worklist: 1519

<u>LAB CASE</u>	<u>ITEM</u>	<u>TASK ID</u>	<u>DESCRIPTION</u>	
C2017-0145	1	75045	AM 1 Urine Screening EMIT	
C2017-0146	1	75048	AM 1 Urine Screening EMIT	
C2017-0149	1	75057	AM 1 Urine Screening EMIT	
C2017-0154	1	75122	AM 1 Urine Screening EMIT	
C2017-0168	1	75482	AM 1 Urine Screening EMIT	
C2017-0185	2	75526	AM 1 Urine Screening EMIT	
C2017-0197	1	75685	AM 1 Urine Screening EMIT	
M2017-0495	10	75736	AM 1 Urine Screening EMIT	
M2017-0495	11	75806	AM 1 Urine Screening EMIT	
M2017-0495	12	75805	AM 1 Urine Screening EMIT	
M2017-0495	13	75804	AM 1 Urine Screening EMIT	
M2017-0495	14	75803	AM 1 Urine Screening EMIT	
M2017-0495	15	75802	AM 1 Urine Screening EMIT	
M2017-0495	16	75801	AM 1 Urine Screening EMIT	
M2017-0495	17	75800	AM 1 Urine Screening EMIT	
M2017-0495	18	75799	AM 1 Urine Screening EMIT	



AUTO MAINTENANCE

BLANK ROTOR RESULTS

Blank date/time: 2/2/2017 2:06:54 PM

Wavelength	Cuvette AV	Cuvette SD	Errors	Lamp AV	Lamp SD	Errors
340	0.4457	0.0186		1.9867	0.0007	
415	0.3423	0.0138		0.8515	0.0005	
505	0.2834	0.0126		0.2796	0.0004	
546	0.2669	0.0123		0.2025	0.0004	
570	0.2701	0.0120		0.2159	0.0004	
600	0.2624	0.0117		0.1164	0.0010	
660	0.2878	0.0118		0.0447	0.0006	
700	0.2809	0.0116		0.1557	0.0003	

CALIBRATION PLANNING

Test name	Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
+THC 50	DAT Cal	*								
+Amphet 500	DAT Cal	*								
+Benzodiazep300	DAT Cal	*								
+Cocaine 300	DAT Cal	*								
+Opiate 300	DAT Cal	*								

MAINTENANCE PLANNING

Maintenance procedure	0	1	2	3	4	5	6	7	>7 days valid
NEEDLE RINSE								*	
SYSTEM CLEAN									*

EXPIRING REAGENTS

Reagent name	0	1	2	3	4	5	6	7	>7 days valid

EXPIRING CONTROLS

Control name	0	1	2	3	4	5	6	7	>7 days valid
Level 0									*
Level 5									*
S2E									*
S2S									*

EXPIRING CALIBRATORS

Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
Acetaminophen									*
Salicylate Cal									*
sTOX Cal									*
sTCA Cal									*
ETOH Cal									*
DAT Cal									*
Ecstasy Cal 2									*
Ecstasy Cal 3									*
LSD Cal									*
qEcstasyCal 300									*
qEcstasyCal 500									*
qLSD Cal									*
cR Cal									*
sG Cal									*
pH Cal									*
cH Cal									*
nT Cal									*
oX Cal									*

AUTO MAINTENANCE**BLANK ROTOR RESULTS**

Blank date/time: 2/2/2017 1:41:15 PM

Wavelength	Cuvette AV	Cuvette SD	Errors	Lamp AV	Lamp SD	Errors
340	369.1565	54.3312	UND SD.ERR	1.9874	0.0008	
415	0.3430	0.0138		0.8517	0.0006	
505	0.2839	0.0125		0.2801	0.0003	
546	0.2673	0.0123		0.2026	0.0004	
570	0.2705	0.0120		0.2160	0.0004	
600	0.2627	0.0117		0.1166	0.0013	
660	0.2881	0.0118		0.0449	0.0007	
700	0.2812	0.0116		0.1555	0.0006	

CALIBRATION PLANNING

Test name	Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
+THC 50	DAT Cal	*								
+Amphet 500	DAT Cal	*								
+Benzodiazep300	DAT Cal	*								
+Cocaine 300	DAT Cal	*								
+Opiate 300	DAT Cal	*								

MAINTENANCE PLANNING

Maintenance procedure	0	1	2	3	4	5	6	7	>7 days valid
NEEDLE RINSE								*	
SYSTEM CLEAN									*

EXPIRING REAGENTS

Reagent name	0	1	2	3	4	5	6	7	>7 days valid

EXPIRING CONTROLS

Control name	0	1	2	3	4	5	6	7	>7 days valid
Level 0									*
Level 5									*
S2E									*
S2S									*

EXPIRING CALIBRATORS

Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
Acetaminophen									*
Salicylate Cal									*
sTOX Cal									*
sTCA Cal									*
ETOH Cal									*
DAT Cal									*
Ecstasy Cal 2									*
Ecstasy Cal 3									*
LSD Cal									*
qEcstasyCal 300									*
qEcstasyCal 500									*
qLSD Cal									*
cR Cal									*
sG Cal									*
pH Cal									*
cH Cal									*
nT Cal									*
oX Cal									*

BLANK/CALIBRATOR INFO

Test name	Reagent blank	Calibrator	
+Amphet 500		Accepted	Yes
		Value	0.3931 dAbs/m
		Meas. date	2/2/2017
		Type	One point - Cut-off
+Benzodiazep300		Accepted	Yes
		Value	0.3935 dAbs/m
		Meas. date	2/2/2017
		Type	One point - Cut-off
+THC 50		Accepted	Yes
		Value	0.3261 dAbs/m
		Meas. date	2/2/2017
		Type	One point - Cut-off
+Cocaine 300		Accepted	Yes
		Value	0.3661 dAbs/m
		Meas. date	2/2/2017
		Type	One point - Cut-off
+Opiate 300		Accepted	Yes
		Value	0.2671 dAbs/m
		Meas. date	2/2/2017
		Type	One point - Cut-off

BW

Measurement date: 2/2/2017 3:08:59 PM

Control name: Level 0

Batch number: 9a508ul-j3

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	negative	0.334 dAbs/m		0.059 dAbs/m	-
+Benzodiazep300	negative	0.284 dAbs/m		0.109 dAbs/m	-
+THC 50	negative	0.281 dAbs/m		0.046 dAbs/m	-
+Cocaine 300	negative	0.249 dAbs/m		0.117 dAbs/m	-
+Opiate 300	negative	0.189 dAbs/m		0.078 dAbs/m	-

Measurement date: 2/2/2017 3:11:13 PM

Control name: Level 5

Batch number: 9A609UL-J4

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.530 dAbs/m		0.137 dAbs/m	+
+Benzodiazep300	POSITIVE	0.480 dAbs/m		0.086 dAbs/m	+
+THC 50	POSITIVE	0.460 dAbs/m		0.134 dAbs/m	+
+Cocaine 300	POSITIVE	0.425 dAbs/m		0.058 dAbs/m	+
+Opiate 300	POSITIVE	0.418 dAbs/m		0.151 dAbs/m	+

Measurement date: 2/2/2017 3:13:01 PM

Control name: S2E

Batch number: 72520

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Benzodiazep300	POSITIVE	0.466 dAbs/m		0.072 dAbs/m	+
+THC 50	POSITIVE	0.432 dAbs/m		0.106 dAbs/m	+
+Cocaine 300	POSITIVE	0.423 dAbs/m		0.057 dAbs/m	+
+Opiate 300	POSITIVE	0.346 dAbs/m		0.079 dAbs/m	+

Measurement date: 2/2/2017 3:22:00 PM

Control name: S2S

Batch number: 60100

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.522 dAbs/m		0.129 dAbs/m	+

WORKLIST

Pos	Type	Rep	Sample data	Tests requested		
1	Control	1	ID: Level 0	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
2	Control	1	ID: Level 5	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
3	Control	1	ID: S2E	1 +Benzodiazep300 4 +Opiate 300	2 +THC 50	3 +Cocaine 300
4	Control	1	ID: S2S	1 +Amphet 500		
5	Sample	1	ID: brc3-68430 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
6	Sample	1	ID: brs3-67140 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
7	Sample	1	ID: neg - 121616 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
8	Sample	1	ID: c2017-0145-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
9	Sample	1	ID: c2017-0146-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
10	Sample	1	ID: c2017-0149-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
11	Sample	1	ID: c2017-0154-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
12	Sample	1	ID: c2017-0185-2 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
13	Sample	1	ID: c2017-0197-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
14	Sample	1	ID: m20170495-10 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
15	Sample	1	ID: m20170495-12 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
16	Sample	1	ID: m20170495-13 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
17	Sample	1	ID: m20170495-14 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
18	Sample	1	ID: c2017-0168-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
19	Sample	1	ID: m20170495-15 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
20	Sample	1	ID: m20170495-16 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
21	Sample	1	ID: m20170495-17 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
22	Sample	1	ID: m20170495-18 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
23	Sample	1	ID: m20170495-11 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen
Analytical Method 1.1

Laboratory Case No: neg - 121616
Date of Analysis: 2/2/2017
Time of Analysis: 3:27:25 PM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	negative 0.336	0.393
Benzodiazepines (300 ng/mL)	negative 0.283	0.394
Cannabinoids (50 ng/mL)	negative 0.281	0.326
Cocaine Metabolite (300 ng/mL)	negative 0.247	0.366
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	negative 0.188	0.267

VIVA-Jr Instrument ID: 8-1324



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen
Analytical Method 1.1

Laboratory Case No: brc3-68430
Date of Analysis: 2/2/2017
Time of Analysis: 3:22:55 PM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.413	0.393
Benzodiazepines (300 ng/mL)	POSITIVE 0.629	0.394
Cannabinoids (50 ng/mL)	negative 0.293	0.326
Cocaine Metabolite (300 ng/mL)	negative 0.304	0.366
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.439	0.267

VIVA-Jr Instrument ID: 8-1324



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen
Analytical Method 1.1

Laboratory Case No: brs3-67140
Date of Analysis: 2/2/2017
Time of Analysis: 3:25:10 PM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.491	0.393
Benzodiazepines (300 ng/mL)	POSITIVE 0.632	0.394
Cannabinoids (50 ng/mL)	POSITIVE 0.423	0.326
Cocaine Metabolite (300 ng/mL)	POSITIVE 0.387	0.366
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.353	0.267

VIVA-Jr Instrument ID: 8-1324