














**Worklist: 1730**

<u>LAB CASE</u>	<u>ITEM</u>	<u>TASK ID</u>	<u>DESCRIPTION</u>	
C2017-0890	1	84365	AM 1 Urine Screening EMIT	
C2017-0901	1	84505	AM 1 Urine Screening EMIT	
C2017-0936	1	84783	AM 1 Urine Screening EMIT	
C2017-0937	1	84787	AM 1 Urine Screening EMIT	
C2017-0940	1	84814	AM 1 Urine Screening EMIT	
C2017-0950	1	84902	AM 1 Urine Screening EMIT	
C2017-0969	1	85157	AM 1 Urine Screening EMIT	
M2017-2126	3	84606	AM 1 Urine Screening EMIT	
M2017-2129	3	84613	AM 1 Urine Screening EMIT	
M2017-2130	3	84603	AM 1 Urine Screening EMIT	
M2017-2131	3	84616	AM 1 Urine Screening EMIT	
M2017-2204	2	84934	AM 1 Urine Screening EMIT	



# AUTO MAINTENANCE

## BLANK ROTOR RESULTS

Blank date/time: 5/22/2017 1:22:49 PM

Wavelength	Cuvette AV	Cuvette SD	Errors	Lamp AV	Lamp SD	Errors
340	0.3953	0.0167		1.9453	0.0006	
415	0.3148	0.0064		0.7899	0.0006	
505	0.2572	0.0056		0.2225	0.0003	
546	0.2401	0.0054		0.1519	0.0004	
570	0.2451	0.0054		0.1622	0.0005	
600	0.2663	0.0052		0.0466	0.0012	
660	0.2655	0.0052		-0.0021	0.0007	
700	0.2909	0.0050		0.0642	0.0006	

## CALIBRATION PLANNING

Test name	Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
+THC 50	DAT Cal	*								
+Amphet 500	DAT Cal	*								
+Benzodiazep300	DAT Cal	*								
+Cocaine 300	DAT Cal	*								
+Opiate 300	DAT Cal	*								

## MAINTENANCE PLANNING

Maintenance procedure	0	1	2	3	4	5	6	7	>7 days valid
NEEDLE RINSE								*	
SYSTEM CLEAN									*

## EXPIRING REAGENTS

Reagent name	0	1	2	3	4	5	6	7	>7 days valid

## EXPIRING CONTROLS

Control name	0	1	2	3	4	5	6	7	>7 days valid
Level 0									*
Level 5									*
S2E									*
S2S									*

## EXPIRING CALIBRATORS

Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
Acetaminophen									*
Salicylate Cal									*
sTOX Cal									*
sTCA Cal									*
ETOH Cal									*
DAT Cal									*
Ecstasy Cal 2									*
Ecstasy Cal 3									*
LSD Cal									*
qEcstasyCal 300									*
qEcstasyCal 500									*
qLSD Cal									*
cR Cal									*
sG Cal									*
pH Cal									*
cH Cal									*
nT Cal									*
oX Cal									*

Measurement date: 5/22/2017 1:45:37 PM

Control name: Level 0

Batch number: 9a508ul-j3

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	negative	0.337	dAbs/m	0.043	dAbs/m -
+Benzodiazep300	negative	0.277	dAbs/m	0.124	dAbs/m -
+THC 50	negative	0.278	dAbs/m	0.046	dAbs/m -
+Cocaine 300	negative	0.246	dAbs/m	0.117	dAbs/m -
+Opiate 300	negative	0.187	dAbs/m	0.082	dAbs/m -

Measurement date: 5/22/2017 1:47:52 PM

Control name: Level 5

Batch number: 9A609UL-J4

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.540	dAbs/m	0.161	dAbs/m +
+Benzodiazep300	POSITIVE	0.495	dAbs/m	0.094	dAbs/m +
+THC 50	POSITIVE	0.467	dAbs/m	0.143	dAbs/m +
+Cocaine 300	POSITIVE	0.418	dAbs/m	0.055	dAbs/m +
+Opiate 300	POSITIVE	0.424	dAbs/m	0.155	dAbs/m +

Measurement date: 5/22/2017 1:49:40 PM

Control name: S2E

Batch number: 72520

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Benzodiazep300	POSITIVE	0.433	dAbs/m	0.032	dAbs/m +
+THC 50	POSITIVE	0.366	dAbs/m	0.043	dAbs/m +
+Cocaine 300	POSITIVE	0.389	dAbs/m	0.026	dAbs/m +
+Opiate 300	POSITIVE	0.330	dAbs/m	0.061	dAbs/m +

Measurement date: 5/22/2017 1:58:39 PM

Control name: S2S

Batch number: 60100

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.463	dAbs/m	0.084	dAbs/m +

## WORKLIST

Pos	Type	Rep	Sample data	Tests requested
5	Sample	1	ID: brc3-68440 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
6	Sample	1	ID: brs3-67140 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
7	Sample	1	ID: neg-42617 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
8	Sample	1	ID: c2017-0890-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
9	Sample	1	ID: c2017-0901-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
10	Sample	1	ID: c2017-0936-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
11	Sample	1	ID: c2017-0937-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
12	Sample	1	ID: c2017-0940-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
13	Sample	1	ID: c2017-0950-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
14	Sample	1	ID: m2017-2126-3 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
15	Sample	1	ID: m2017-2129-3 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
16	Sample	1	ID: m2017-2130-3 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
17	Sample	1	ID: m2017-2131-3 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
18	Sample	1	ID: m2017-2204-2 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
19	Sample	1	ID: c2017-0969-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50



## Idaho State Police Forensic Services

**EMIT Urine Drugs-of-Abuse Screen**  
Analytical Method 1.1

Laboratory Case No: neg-42617  
Date of Analysis: 5/22/2017  
Time of Analysis: 2:26:34 PM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	negative 0.338	0.380
Benzodiazepines (300 ng/mL)	negative 0.281	0.401
Cannabinoids (50 ng/mL)	negative 0.284	0.324
Cocaine Metabolite (300 ng/mL)	negative 0.245	0.363
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	negative 0.188	0.269

VIVA-Jr Instrument ID: 8-1324



## Idaho State Police Forensic Services

**EMIT Urine Drugs-of-Abuse Screen**  
Analytical Method 1.1

Laboratory Case No: brc3-68440  
Date of Analysis: 5/22/2017  
Time of Analysis: 2:23:23 PM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.409	0.380
Benzodiazepines (300 ng/mL)	POSITIVE 0.631	0.401
Cannabinoids (50 ng/mL)	negative 0.294	0.324
Cocaine Metabolite (300 ng/mL)	negative 0.290	0.363
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.442	0.269

VIVA-Jr Instrument ID: 8-1324



## Idaho State Police Forensic Services

**EMIT Urine Drugs-of-Abuse Screen**  
Analytical Method 1.1

Laboratory Case No: brs3-67140  
Date of Analysis: 5/22/2017  
Time of Analysis: 2:24:18 PM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.510	0.380
Benzodiazepines (300 ng/mL)	POSITIVE 0.621	0.401
Cannabinoids (50 ng/mL)	POSITIVE 0.411	0.324
Cocaine Metabolite (300 ng/mL)	POSITIVE 0.389	0.363
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.345	0.269

VIVA-Jr Instrument ID: 8-1324