



















**Worklist: 1748**

*Bylee*

<u>LAB CASE</u>	<u>ITEM</u>	<u>TASK ID</u>	<u>DESCRIPTION</u>	
C2017-0976	1	85270	AM 1 Urine Screening EMIT	
C2017-0979	1	85286	AM 1 Urine Screening EMIT	
C2017-0988	1	85305	AM 1 Urine Screening EMIT	
C2017-1005	3	85461	AM 1 Urine Screening EMIT	
C2017-1006	1	85464	AM 1 Urine Screening EMIT	
C2017-1011	1	85484	AM 1 Urine Screening EMIT	
C2017-1017	1	85514	AM 1 Urine Screening EMIT	
C2017-1018	1	85519	AM 1 Urine Screening EMIT	
C2017-1020	1	85651	AM 1 Urine Screening EMIT	
C2017-1021	1	85656	AM 1 Urine Screening EMIT	
C2017-1022	1	85748	AM 1 Urine Screening EMIT	
C2017-1026	1	85804	AM 1 Urine Screening EMIT	
C2017-1052	2	86260	AM 1 Urine Screening EMIT	
C2017-1074	1	86101	AM 1 Urine Screening EMIT	
M2017-2208	4	86234	AM 1 Urine Screening EMIT	
M2017-2290	3	86083	AM 1 Urine Screening EMIT	
M2017-2407	2	86097	AM 1 Urine Screening EMIT	
M2017-2411	2	85672	AM 1 Urine Screening EMIT	



# AUTO MAINTENANCE

## BLANK ROTOR RESULTS

Blank date/time: 6/6/2017 4:15:28 AM

Wavelength	Cuvette AV	Cuvette SD	Errors	Lamp AV	Lamp SD	Errors
340	0.4007	0.0066		1.9423	0.0007	
415	0.3516	0.0056		0.7838	0.0007	
505	0.2582	0.0054		0.2180	0.0005	
546	0.2411	0.0053		0.1478	0.0005	
570	0.2459	0.0053		0.1589	0.0005	
600	0.2672	0.0052		0.0327	0.0014	
660	0.2660	0.0052		-0.0061	0.0008	
700	0.2913	0.0050		0.0614	0.0004	

## CALIBRATION PLANNING

Test name	Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
+THC 50	DAT Cal	*								
+Amphet 500	DAT Cal	*								
+Benzodiazep300	DAT Cal	*								
+Cocaine 300	DAT Cal	*								
+Opiate 300	DAT Cal	*								

## MAINTENANCE PLANNING

Maintenance procedure	0	1	2	3	4	5	6	7	>7 days valid
NEEDLE RINSE								*	
SYSTEM CLEAN									*

## EXPIRING REAGENTS

Reagent name	0	1	2	3	4	5	6	7	>7 days valid

## EXPIRING CONTROLS

Control name	0	1	2	3	4	5	6	7	>7 days valid
Level 0									*
Level 5									*
S2E									*
S2S									*

\* OK added new control expires 5-31-19

## EXPIRING CALIBRATORS

Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
Acetaminophen									*
Salicylate Cal									*
sTOX Cal									*
sTCA Cal									*
ETOH Cal									*
DAT Cal									*
Ecstasy Cal 2									*
Ecstasy Cal 3									*
LSD Cal									*
qEcstasyCal 300									*
qEcstasyCal 500									*
qLSD Cal									*
cR Cal									*
sG Cal									*
pH Cal									*
cH Cal									*
nT Cal									*
oX Cal									*

Measurement date: 6/6/2017 7:46:26 AM

Control name: Level 0

Batch number: 9a508ul-j3

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	negative	0.333	dAbs/m	0.046	dAbs/m -
+Benzodiazep300	negative	0.281	dAbs/m	0.120	dAbs/m -
+THC 50	negative	0.280	dAbs/m	0.044	dAbs/m -
+Cocaine 300	negative	0.245	dAbs/m	0.118	dAbs/m -
+Opiate 300	negative	0.191	dAbs/m	0.078	dAbs/m -

Measurement date: 6/6/2017 7:48:41 AM

Control name: Level 5

Batch number: 9A609UL-J4

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.543	dAbs/m	0.163	dAbs/m +
+Benzodiazep300	POSITIVE	0.488	dAbs/m	0.086	dAbs/m +
+THC 50	POSITIVE	0.473	dAbs/m	0.150	dAbs/m +
+Cocaine 300	POSITIVE	0.417	dAbs/m	0.054	dAbs/m +
+Opiate 300	POSITIVE	0.429	dAbs/m	0.160	dAbs/m +

Measurement date: 6/6/2017 7:50:30 AM

Control name: S2E

Batch number: 72520

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Benzodiazep300	POSITIVE	0.439	dAbs/m	0.038	dAbs/m +
+THC 50	POSITIVE	0.355	dAbs/m	0.032	dAbs/m +
+Cocaine 300	POSITIVE	0.389	dAbs/m	0.026	dAbs/m +
+Opiate 300	POSITIVE	0.338	dAbs/m	0.070	dAbs/m +

Measurement date: 6/6/2017 7:59:28 AM

Control name: S2S

Batch number: 60130

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.420	dAbs/m	0.040	dAbs/m +



## Idaho State Police Forensic Services

**EMIT Urine Drugs-of-Abuse Screen**  
Analytical Method 1.1

Laboratory Case No: brc3 68440  
Date of Analysis: 6/6/2017  
Time of Analysis: 8:43:06 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.412	0.380
Benzodiazepines (300 ng/mL)	POSITIVE 0.609	0.401
Cannabinoids (50 ng/mL)	negative 0.287	0.324
Cocaine Metabolite (300 ng/mL)	negative 0.285	0.363
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.444	0.269

VIVA-Jr Instrument ID: 8-1324



## Idaho State Police Forensic Services

**EMIT Urine Drugs-of-Abuse Screen**  
Analytical Method 1.1

Laboratory Case No: brs3 67140  
Date of Analysis: 6/6/2017  
Time of Analysis: 8:45:21 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.509	0.380
Benzodiazepines (300 ng/mL)	POSITIVE 0.618	0.401
Cannabinoids (50 ng/mL)	POSITIVE 0.392	0.324
Cocaine Metabolite (300 ng/mL)	POSITIVE 0.382	0.363
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.353	0.269

VIVA-Jr Instrument ID: 8-1324



## Idaho State Police Forensic Services

**EMIT Urine Drugs-of-Abuse Screen**  
Analytical Method 1.1

Laboratory Case No: neg 42617  
Date of Analysis: 6/6/2017  
Time of Analysis: 8:54:48 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	negative 0.337	0.380
Benzodiazepines (300 ng/mL)	negative 0.283	0.401
Cannabinoids (50 ng/mL)	negative 0.284	0.324
Cocaine Metabolite (300 ng/mL)	negative 0.244	0.363
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	negative 0.188	0.269

VIVA-Jr Instrument ID: 8-1324