




















9/19/2017

Worklist: 1897

<u>LAB_CASE</u>	<u>ITEM</u>	<u>TASK_ID</u>	<u>DESCRIPTION</u>	
C2017-1829	1	95186	AM 1 Urine Screening EMIT	
C2017-1830	1	95187	AM 1 Urine Screening EMIT	
C2017-1836	1	94737	AM 1 Urine Screening EMIT	
C2017-1849	1	94759	AM 1 Urine Screening EMIT	
C2017-1861	1	94855	AM 1 Urine Screening EMIT	
C2017-1867	1	94869	AM 1 Urine Screening EMIT	
M2017-4110	6	95173	AM 1 Urine Screening EMIT	
P2017-2019	1	95197	AM 1 Urine Screening EMIT	
P2017-2020	1	95196	AM 1 Urine Screening EMIT	
P2017-2021	1	95195	AM 1 Urine Screening EMIT	
P2017-2083	1	95194	AM 1 Urine Screening EMIT	
P2017-2086	1	95193	AM 1 Urine Screening EMIT	
P2017-2112	1	95192	AM 1 Urine Screening EMIT	
P2017-2155	1	95191	AM 1 Urine Screening EMIT	
P2017-2162	1	95190	AM 1 Urine Screening EMIT	
P2017-2163	1	95189	AM 1 Urine Screening EMIT	
P2017-2164	1	95188	AM 1 Urine Screening EMIT	



AUTO MAINTENANCE

BLANK ROTOR RESULTS

Blank date/time: 9/18/2017 4:15:31 AM

Wavelength	Cuvette AV	Cuvette SD	Errors	Lamp AV	Lamp SD	Errors
340	0.3760	0.0067		1.9714	0.0009	
415	0.3219	0.0056		0.8006	0.0006	
505	0.2283	0.0052		0.2350	0.0004	
546	0.2115	0.0051		0.1669	0.0004	
570	0.2180	0.0051		0.1706	0.0004	
600	0.2404	0.0050		0.0484	0.0014	
660	0.2379	0.0049		0.0092	0.0006	
700	0.2640	0.0048		0.0761	0.0003	

CALIBRATION PLANNING

Test name	Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
+THC 50	DAT Cal	*								
+Amphet 500	DAT Cal	*								
+Benzodiazep300	DAT Cal	*								
+Cocaine 300	DAT Cal	*								
+Opiate 300	DAT Cal	*								

MAINTENANCE PLANNING

Maintenance procedure	0	1	2	3	4	5	6	7	>7 days valid
NEEDLE RINSE						*			
SYSTEM CLEAN									*

EXPIRING REAGENTS

Reagent name	0	1	2	3	4	5	6	7	>7 days valid

EXPIRING CONTROLS

Control name	0	1	2	3	4	5	6	7	>7 days valid
Level 0									*
Level 5									*
S2E									*
S2S									*

EXPIRING CALIBRATORS

Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
Acetaminophen									*
Salicylate Cal									*
sTOX Cal									*
sTCA Cal									*
ETOH Cal									*
DAT Cal									*
Ecstasy Cal 2									*
Ecstasy Cal 3									*
LSD Cal									*
qEcstasyCal 300									*
qEcstasyCal 500									*
qLSD Cal									*
cR Cal									*
sG Cal									*
pH Cal									*
cH Cal									*
nT Cal									*
oX Cal									*

BLANK/CALIBRATOR INFO

Test name	Reagent blank	Calibrator	
+Amphet 500		Accepted	Yes
		Value	0.4093 dAbs/m
		Meas. date	9/15/2017
		Type	One point - Cut-off
+Benzodiazep300		Accepted	Yes
		Value	0.3947 dAbs/m
		Meas. date	9/15/2017
		Type	One point - Cut-off
+THC 50		Accepted	Yes
		Value	0.3303 dAbs/m
		Meas. date	9/15/2017
		Type	One point - Cut-off
+Cocaine 300		Accepted	Yes
		Value	0.3729 dAbs/m
		Meas. date	9/15/2017
		Type	One point - Cut-off
+Opiate 300		Accepted	Yes
		Value	0.2687 dAbs/m
		Meas. date	9/15/2017
		Type	One point - Cut-off

Measurement date: 9/18/2017 7:57:26 AM

Control name: Level 0

Batch number: 9a508ul-k1

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	negative	0.330	dAbs/m	0.080	dAbs/m -
+Benzodiazep300	negative	0.285	dAbs/m	0.110	dAbs/m -
+THC 50	negative	0.290	dAbs/m	0.041	dAbs/m -
+Cocaine 300	negative	0.255	dAbs/m	0.118	dAbs/m -
+Opiate 300	negative	0.193	dAbs/m	0.076	dAbs/m -

Measurement date: 9/18/2017 7:59:41 AM

Control name: Level 5

Batch number: 9A609UL-k1

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.539	dAbs/m	0.130	dAbs/m +
+Benzodiazep300	POSITIVE	0.493	dAbs/m	0.098	dAbs/m +
+THC 50	POSITIVE	0.473	dAbs/m	0.143	dAbs/m +
+Cocaine 300	POSITIVE	0.439	dAbs/m	0.066	dAbs/m +
+Opiate 300	POSITIVE	0.425	dAbs/m	0.156	dAbs/m +

Measurement date: 9/18/2017 8:01:29 AM

Control name: S2E

Batch number: 72520

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Benzodiazep300	POSITIVE	0.463	dAbs/m	0.068	dAbs/m +
+THC 50	POSITIVE	0.416	dAbs/m	0.085	dAbs/m +
+Cocaine 300	POSITIVE	0.428	dAbs/m	0.055	dAbs/m +
+Opiate 300	POSITIVE	0.362	dAbs/m	0.093	dAbs/m +

Measurement date: 9/18/2017 8:03:16 AM

Control name: S2S

Batch number: 60130

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.497	dAbs/m	0.088	dAbs/m +

WORKLIST

Pos	Type	Rep	Sample data	Tests requested
1	Sample	1	ID: brc3-68430 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
2	Sample	1	ID: brs2s-60130 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
3	Sample	1	ID: neg-9517 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
4	Sample	1	ID: c2017-1829-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
5	Sample	1	ID: c2017-1830-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
6	Sample	1	ID: c2017-1836-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
7	Sample	1	ID: c2017-1849-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
8	Sample	1	ID: c2017-1861-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
9	Sample	1	ID: c2017-1867-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
10	Sample	1	ID: m2017-4110-6 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
11	Sample	1	ID: p2017-2019-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
12	Sample	1	ID: p2017-2020-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
13	Sample	1	ID: p2017-2021-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
14	Sample	1	ID: p2017-2083-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
15	Sample	1	ID: p2017-2086-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
16	Sample	1	ID: p2017-2112-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
17	Sample	1	ID: p2017-2155-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
18	Sample	1	ID: p2017-2162-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
19	Sample	1	ID: p2017-2163-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50
20	Sample	1	ID: p2017-2164-1 Name:	1 +Amphet 500 4 +Cocaine 300 2 +Benzodiazep300 5 +Opiate 300 3 +THC 50



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen
Analytical Method 1.1

Laboratory Case No: neg-9517
Date of Analysis: 9/18/2017
Time of Analysis: 2:50:00 PM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	negative 0.334	0.409
Benzodiazepines (300 ng/mL)	negative 0.282	0.395
Cannabinoids (50 ng/mL)	negative 0.290	0.330
Cocaine Metabolite (300 ng/mL)	negative 0.251	0.373
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	negative 0.193	0.269

VIVA-Jr Instrument ID: 8-1324



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen
Analytical Method 1.1

Laboratory Case No: brc3-68430
Date of Analysis: 9/18/2017
Time of Analysis: 2:38:18 PM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.434	0.409
Benzodiazepines (300 ng/mL)	POSITIVE 0.625	0.395
Cannabinoids (50 ng/mL)	negative 0.296	0.330
Cocaine Metabolite (300 ng/mL)	negative 0.299	0.373
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.453	0.269

VIVA-Jr Instrument ID: 8-1324



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen
Analytical Method 1.1

Laboratory Case No: brs2s-60130
Date of Analysis: 9/18/2017
Time of Analysis: 2:40:34 PM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.449	0.409
Benzodiazepines (300 ng/mL)	negative 0.281	0.395
Cannabinoids (50 ng/mL)	POSITIVE 0.346	0.330
Cocaine Metabolite (300 ng/mL)	negative 0.343 elevated - OK baw 9/21/17	0.373
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.408	0.269

VIVA-Jr Instrument ID: 8-1324