



File : I:\Instrument Data\Pocatello\Major Mass Spec\CDS\2018\am 2 o  
... ptional gluc. step infreq. used test pre-run 082318\Negative  
... Control.D  
Instrument : Major Mass Spec  
Operator : ISP\datastor  
Acquired : 23 Aug 2018 13:50 using AcqMethod UTOX GEN DRUG 6-2017extended.M  
Sample Name: Negative Control


Because the optional glucuronidation step is not done frequently, an extraction with just the controls was done and analyzed before the sample extraction was done. *cs*



**REVIEWED**



By Brittany Wylie at 11:20 am, Sep 06, 2018

File :I:\Instrument Data\Pocatello\Major Mass Spec\CDS\2018\am 2 o  
... ptional gluc. step infreq. used test pre-run 082318\Positive  
... Control.D  
Instrument : Major Mass Spec  
Operator : ISP\datastor  
Acquired : 23 Aug 2018 14:16 using AcqMethod UTOX GEN DRUG 6-2017extended.M  
Sample Name: Positive Control





File : I:\Instrument Data\Pocatello\Major Mass Spec\CDS\2018\am 2 o  
... ptional gluc. step infreq. used test pre-run 082318\Positive  
... Control.D  
Instrument : Major Mass Spec  
Operator : ISP\datastor  
Acquired : 23 Aug 2018 14:16 using AcqMethod UTOX GEN DRUG 6-2017extended.M  
Sample Name: Positive Control

CS





File : I:\Instrument Data\Pocatello\Major Mass Spec\CDS\2018\am 2 o  
... ptional gluc. step infreq. used test pre-run 082318\Positive  
... Control.D  
Instrument : Major Mass Spec  
Operator : ISP\datastor  
Acquired : 23 Aug 2018 14:16 using AcqMethod UTOX GEN DRUG 6-2017extended.M  
Sample Name: Positive Control

CS





File : I:\Instrument Data\Pocatello\Major Mass Spec\CDS\2018\am 2 o  
... ptional gluc. step infreq. used test pre-run 082318\Positive  
... Control.D  
Instrument : Major Mass Spec  
Operator : ISP\datastor  
Acquired : 23 Aug 2018 14:16 using AcqMethod UTOX GEN DRUG 6-2017extended.M  
Sample Name: Positive Control

CS





File : I:\Instrument Data\Pocatello\Major Mass Spec\CDS\2018\am 2 o  
... ptional gluc. step infreq. used test pre-run 082318\Positive  
... Control.D  
Instrument : Major Mass Spec  
Operator : ISP\datastor  
Acquired : 23 Aug 2018 14:16 using AcqMethod UTOX GEN DRUG 6-2017extended.M  
Sample Name: Positive Control

CS




File : I:\Instrument Data\Pocatello\Major Mass Spec\CDS\2018\am 2 o  
... ptional gluc. step infreq. used test pre-run 082318\Positive  
... Control.D  
Instrument : Major Mass Spec  
Operator : ISP\datastor  
Acquired : 23 Aug 2018 14:16 using AcqMethod UTOX GEN DRUG 6-2017extended.M  
Sample Name: Positive Control

CS




File :I:\Instrument Data\Pocatello\Major Mass Spec\CDS\2018\am 2 o  
... ptional gluc. step infreq. used test pre-run 082318\Glucuron  
... ide Negative Control.D  
Instrument : Major Mass Spec  
Operator : ISP\datastor  
Acquired : 23 Aug 2018 15:08 using AcqMethod UTOX GEN DRUG 6-2017extended.M  
Sample Name: Negative Control







File :I:\Instrument Data\Pocatello\Major Mass Spec\CDS\2018\am 2 o  
... ptional gluc. step infreq. used test pre-run 082318\Glucuron  
... ide Negative Control.D  
Instrument : Major Mass Spec  
Operator : ISP\datastor  
Acquired : 23 Aug 2018 15:08 using AcqMethod UTOX GEN DRUG 6-2017extended.M  
Sample Name: Negative Control






File : I:\Instrument Data\Pocatello\Major Mass Spec\CDS\2018\am 2 o  
... ptional gluc. step infreq. used test pre-run 082318\Glucuron  
... ide Positive Control.D  
Instrument : Major Mass Spec  
Operator : ISP\datastor  
Acquired : 23 Aug 2018 15:35 using AcqMethod UTOX GEN DRUG 6-2017extended.M  
Sample Name: Positive Control

CS




File : I:\Instrument Data\Pocatello\Major Mass Spec\CDS\2018\am 2 w  
... ith optional glucuronidation 082918\Negative Control.D  
Operator : ISP\datastor  
Instrument : Major Mass Spec  
Acquired : 29 Aug 2018 07:55 using AcqMethod UTOX GEN DRUG 6-2017extended.M  
Sample Name: Negative Control  
Misc Info : POC062718; AM #2 With Optional Glucuronidation Step

This is the run that included the case sample. 





File :I:\Instrument Data\Pocatello\Major Mass Spec\CDS\2018\am 2 w  
... ith optional glucuronidation 082918\Positive Control.D  
Operator : ISP\datastor  
Instrument : Major Mass Spec  
Acquired : 29 Aug 2018 11:35 using AcqMethod UTOX GEN DRUG 6-2017extended.M  
Sample Name: Positive Control  
Misc Info : POC062718 + WS082318; AM #2 With Optional Glucuronidation Step





File : I:\Instrument Data\Pocatello\Major Mass Spec\CDS\2018\am 2 w  
... ith optional glucuronidation 082918\Positive Control.D  
Operator : ISP\datastor  
Instrument : Major Mass Spec  
Acquired : 29 Aug 2018 11:35 using AcqMethod UTOX GEN DRUG 6-2017extended.M  
Sample Name: Positive Control  
Misc Info : POC062718 + WS082318; AM #2 With Optional Glucuronidation Step

CS





File : I:\Instrument Data\Pocatello\Major Mass Spec\CDS\2018\am 2 w  
... ith optional glucuronidation 082918\Positive Control.D  
Operator : ISP\datastor  
Instrument : Major Mass Spec  
Acquired : 29 Aug 2018 11:35 using AcqMethod UTOX GEN DRUG 6-2017extended.M  
Sample Name: Positive Control  
Misc Info : POC062718 + WS082318; AM #2 With Optional Glucuronidation Step

CS





File : I:\Instrument Data\Pocatello\Major Mass Spec\CDS\2018\am 2 w  
... ith optional glucuronidation 082918\Positive Control.D  
Operator : ISP\datastor  
Instrument : Major Mass Spec  
Acquired : 29 Aug 2018 11:35 using AcqMethod UTOX GEN DRUG 6-2017extended.M  
Sample Name: Positive Control  
Misc Info : POC062718 + WS082318; AM #2 With Optional Glucuronidation Step

CS



File : I:\Instrument Data\Pocatello\Major Mass Spec\CDS\2018\am 2 w  
... ith optional glucuronidation 082918\Positive Control.D  
Operator : ISP\datastor  
Instrument : Major Mass Spec  
Acquired : 29 Aug 2018 11:35 using AcqMethod UTOX GEN DRUG 6-2017extended.M  
Sample Name: Positive Control  
Misc Info : POC062718 + WS082318; AM #2 With Optional Glucuronidation Step



CS







File : I:\Instrument Data\Pocatello\Major Mass Spec\CDS\2018\am 2 w  
... ith optional glucuronidation 082918\Positive Control.D  
Operator : ISP\datastor  
Instrument : Major Mass Spec  
Acquired : 29 Aug 2018 11:35 using AcqMethod UTOX GEN DRUG 6-2017extended.M  
Sample Name: Positive Control  
Misc Info : POC062718 + WS082318; AM #2 With Optional Glucuronidation Step

CS




File : I:\Instrument Data\Pocatello\Major Mass Spec\CDS\2018\am 2 w  
... ith optional glucuronidation 082918\Gluc NEG.D  
Operator : ISP\datastor  
Instrument : Major Mass Spec  
Acquired : 29 Aug 2018 12:27 using AcqMethod UTOX GEN DRUG 6-2017extended.M  
Sample Name: Glucuronide Negative Control  
Misc Info : POC062718 + Morphine Glucuronide (Cerilliant Lot No.: 3523306-A); AM #2 With Optional Gluc


CS



File : I:\Instrument Data\Pocatello\Major Mass Spec\CDS\2018\am 2 w  
... ith optional glucuronidation 082918\Gluc NEG.D  
Operator : ISP\datastor  
Instrument : Major Mass Spec  
Acquired : 29 Aug 2018 12:27 using AcqMethod UTOX GEN DRUG 6-2017extended.M  
Sample Name: Glucuronide Negative Control  
Misc Info : POC062718 + Morphine Glucuronide (Cerilliant Lot No.: 3523306-A); AM #2 With Optional



File :I:\Instrument Data\Pocatello\Major Mass Spec\CDS\2018\am 2 w  
... ith optional glucuronidation 082918\Gluc POS.D  
Operator : ISP\datastor  
Instrument : Major Mass Spec  
Acquired : 29 Aug 2018 12:53 using AcqMethod UTOX GEN DRUG 6-2017extended.M  
Sample Name: Glucuronide Positive Control  
Misc Info : POC062718 + Morphine Glucuronide (CerilliantLot No.: 3523306-A); AM #2 With Optional



File : I:\Instrument Data\Pocatello\Major Mass Spec\CDS\2018\am 2 w  
... ith optional glucuronidation 082918\Gluc POS.D  
Operator : ISP\datastor  
Instrument : Major Mass Spec  
Acquired : 29 Aug 2018 12:53 using AcqMethod UTOX GEN DRUG 6-2017extended.M  
Sample Name: Glucuronide Positive Control  
Misc Info : POC062718 + Morphine Glucuronide (CerilliantLot No.: 3523306-A); AM #2 With Optional Gluc

CS

