




Idaho State Police Forensic Services

CERTIFICATE OF APPROVAL

The Idaho State Police Forensic Services (ISPFS) hereby certifies and approves Alcohol Simulator Solution **Lot Number 12804** (a product manufactured by RepCo Marketing Inc.) to be used to conduct performance verification checks within the State of Idaho in accordance with the analytical methods, policies and/or procedures promulgated by the Department governing breath alcohol examinations. ISPFS also approves of the manufacturer of this solution (RepCo Marketing Inc.) to provide Alcohol Simulator Solution **Lot Number 12804** in the State of Idaho. This lot has a target value of 0.080 with a range of 0.072 to 0.088 grams of ethyl alcohol/210 liters of vapor.

The expiration date for this lot number is on September 26th, 2014 at 11:59 PM.

11-7-2012
Date


Breath Alcohol Discipline Leader

STATE OF IDAHO)
) ss.
County of Kootenai)

On this 7th day of November, in the year 2012 before me, Lynn A. Higdem, a notary public, Jeremy Johnston personally appeared, known to me to be the person whose name is subscribed to the within instrument as a Forensic Scientist for the Idaho State Police Forensic Services, and acknowledged to me that he executed the same as such Scientist.

Lynn A. Higdem Notary Public
My Commission Expires: 9/8/2016

