




Idaho State Police Forensic Services

CERTIFICATE OF ANALYSIS/APPROVAL

The Idaho State Police Forensic Services (ISPFS) hereby certifies and approves Alcohol Simulator Solution **Lot Number 19100** (a product manufactured by GUTH Laboratories Inc.) to be used to conduct performance verification checks within the State of Idaho in accordance with the analytical methods, policies and/or procedures promulgated by the Department governing breath alcohol examinations. ISPFS also approves of the manufacturer of this solution (GUTH Laboratories Inc.) to provide Alcohol Simulator Solution **Lot Number 19100** in the State of Idaho. This lot has a target value of 0.080 with a range of 0.072 to 0.088 grams of ethyl alcohol/210 liters of vapor.


The expiration date for this lot number is on March 26th, 2021 at 11:59 PM.


8-30-19
Date


Volatiles Analysis Discipline Leader

STATE OF IDAHO)
) ss.
County of Kootenai)

On this 30th day of August, in the year 2019, before me, Lynn A. Higdem, a notary public, Jeremy Johnston personally appeared, known to me to be the person whose name is subscribed to the within instrument as a Forensic Scientist for the Idaho State Police Forensic Services, and acknowledged to me that he executed the same as such Scientist.

 Notary Public
My Commission Expires: Sept. 8, 2022





GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **19100** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **March 27, 2019**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.0971%** (w/vol) ethyl alcohol. The expiration date for this lot number is **March 26, 2021** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.080 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN02221601 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.