

Director

## Idaho State Police

## Service Since 1939



Brad Little Governor

To: Idaho Law Enforcement

From: Matthew Gamette, ISP Forensic Services Laboratory System Director

**Subject:** Lifeloc FC20 (non-BT version) End of Service Notification

**Date:** June 28, 2022

With manufacturer supply of repair and service parts for the original (non-bluetooth) version of the Lifeloc FC20 becoming obsolete, the original (non-bluetooth) FC20 version will no longer be available for purchase from Lifeloc Corporation. Further, service on the original (non-bluetooth) version of the FC20 will be limited to the parts that Lifeloc Inc. has on hand. The "End of Service" for the original (non-bluetooth) version of the Lifeloc FC20 is considered to be immediate.

ISPFS will continue to calibrate and support the original (non-bluetooth) FC20 model through ISPFS as long as possible. If factory service is required for an original (non-bluetooth) version Lifeloc FC20, please call Lifeloc Inc. and inquire about the availability of the service parts prior to sending the unit to the manufacturer. Lifeloc is offering a trade in program for their FC20 (non-bluetooth) units in order to help facilitate and ease the financial burden of upgrading to the FC20-BT. The BT units are compatible with the current EasyCal as well as the printers.

This End of Service notification <u>only applies to the factory service and parts available</u> for the FC20 (non-BT version). These instruments are still approved for evidentiary collection of breath samples so long as they continue to pass all the requirements set out in the ISP Standard Operating Procedure and IDAPA.

Lifeloc can be reached at 1-800-722-4872 for questions about FC20 service as well as the specifics of their trade-in program.

Technical questions about this bulletin should be directed to Alcohol Discipline Leader Jeremy Johnston @isp.idaho.gov or 208-209-8706.

700 S. Stratford Drive • Meridian, Idaho 83642-6202