




Idaho State Police Forensic Services

CERTIFICATE OF APPROVAL

The Idaho State Police Forensic Services (ISPFS) hereby certifies and approves Alcohol Simulator Solution **Lot Number 11802** (a product manufactured by RepCo Marketing Inc.) to be used to conduct performance verification checks within the State of Idaho in accordance with the analytical methods, policies and/or procedures promulgated by the Department governing breath alcohol examinations. ISPFS also approves of the manufacturer of this solution (RepCo Marketing Inc.) to provide Alcohol Simulator Solution **Lot Number 11802** in the State of Idaho. This lot has a target value of 0.080 with a range of 0.072 to 0.088 grams of ethyl alcohol/210 liters of vapor.

The expiration date for this lot number is on March 8th, 2013 at 11:59 PM.

4-1-2011
Date


Breath Alcohol Discipline Leader

STATE OF IDAHO)
) ss.
County of Kootenai)

On this 1st day of April, in the year 2011, before me, JoAnn Hutchison, a notary public, Jeremy Johnston personally appeared, known to me to be the person whose name is subscribed to the within instrument as a Forensic Scientist for the Idaho State Police Forensic Services, and acknowledge to me that he executed the same as such Scientist.

 Notary Public
My Commission Expires: 12/5/2015

