



U.S. Department
of Transportation

National Highway
Traffic Safety
Administration

1200 New Jersey Avenue, SE
Washington, DC 20590

FEB 17 2015

Mr. Mark Lary
Director of Regulatory Affairs &
Field Technical Support
Lifeloc Technologies, Inc.
12441 West 49th Avenue, Unit 4
Wheat Ridge, CO 80033

Dear Mr. Lary:

This letter confirms that Lifeloc Technologies, Inc. submitted the FC 10, FC10^{Plus}, FC20BT, FC20, EV30, and Phoenix 6.0BT with the EASYCAL[®] accessory for re-evaluation as evidential breath testers (EBTs) due to software changes to the EASYCAL[®] accessory. These software changes include: an update to the dry gas compensation setting, a new printer driver, relocation of operator selection, increase database sizes, and confirmation of dry gas concentration and lot code upon reboot.


These EBT devices with the EASYCAL[®] accessory have previously been evaluated by the Volpe Transportation Systems Center and found to meet the model specifications for EBT devices, and have been included on the conforming products list (CPL) for EBTs.

A copy of this letter will be provided to the U.S. Department of Transportation Office of Drug and Alcohol Policy and Compliance for their records.

If you require further assistance or have any questions regarding NHTSA's quality assurance program in relation to 49 CFR Part 40, please do not hesitate to contact me at 202-366-1694.

Thank you for advising NHTSA of these changes.

Sincerely,



J. De Carlo Ciccel
Program Analyst

Enclosures

CC: Patrice Kelly, OST
Edward Conde, Volpe