






Worklist: 1500

<u>LAB CASE</u>	<u>ITEM</u>	<u>TASK ID</u>	<u>DESCRIPTION</u>	
C2017-0043	1	73907	AM 1 Urine Screening EMIT	
C2017-0103	1	74542	AM 1 Urine Screening EMIT	
C2017-0104	1	74588	AM 1 Urine Screening EMIT	
C2017-0123	1	74815	AM 1 Urine Screening EMIT	
C2017-0124	1	74830	AM 1 Urine Screening EMIT	
C2017-0130	3	74913	AM 1 Urine Screening EMIT	
C2017-0136	1	74955	AM 1 Urine Screening EMIT	
M2017-0038	1	75001	AM 1 Urine Screening EMIT	
M2017-0314	3	75018	AM 1 Urine Screening EMIT	
P2017-0136	2	74812	AM 1 Urine Screening EMIT	



AUTO MAINTENANCE

BLANK ROTOR RESULTS

Blank date/time: 1/25/2017 11:16:26 AM

Wavelength	Cuvette AV	Cuvette SD	Errors	Lamp AV	Lamp SD	Errors
340	0.4455	0.0187		1.9871	0.0008	
415	0.3424	0.0137		0.8521	0.0006	
505	0.2836	0.0125		0.2799	0.0005	
546	0.2670	0.0123		0.2027	0.0004	
570	0.2703	0.0120		0.2161	0.0004	
600	0.2625	0.0117		0.1163	0.0012	
660	0.2880	0.0117		0.0448	0.0007	
700	0.2810	0.0115		0.1557	0.0004	

CALIBRATION PLANNING

Test name	Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
+THC 50	DAT Cal	*								
+Amphet 500	DAT Cal	*								
+Benzodiazep300	DAT Cal	*								
+Cocaine 300	DAT Cal	*								
+Opiate 300	DAT Cal	*								

MAINTENANCE PLANNING

Maintenance procedure	0	1	2	3	4	5	6	7	>7 days valid
NEEDLE RINSE								*	
SYSTEM CLEAN									*

EXPIRING REAGENTS

Reagent name	0	1	2	3	4	5	6	7	>7 days valid

EXPIRING CONTROLS

Control name	0	1	2	3	4	5	6	7	>7 days valid
Level 0									*
Level 5						*			
S2E									*
S2S									*

EXPIRING CALIBRATORS

Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
Acetaminophen									*
Salicylate Cal									*
sTOX Cal									*
sTCA Cal									*
ETOH Cal									*
DAT Cal									*
Ecstasy Cal 2									*
Ecstasy Cal 3									*
LSD Cal									*
qEcstasyCal 300									*
qEcstasyCal 500									*
qLSD Cal									*
cR Cal									*
sG Cal									*
pH Cal									*
ch Cal									*
nT Cal									*
oX Cal									*

AUTO MAINTENANCE

BLANK ROTOR RESULTS

Blank date/time: 1/25/2017 10:33:28 AM

Wavelength	Cuvette AV	Cuvette SD	Errors	Lamp AV	Lamp SD	Errors
340	376.9977	0.0008	UND	1.9883	0.0008	
415	0.3432	0.0140		0.8530	0.0006	
505	0.2840	0.0128		0.2810	0.0003	
546	0.2674	0.0125		0.2034	0.0004	
570	0.2706	0.0122		0.2164	0.0004	
600	0.2628	0.0119		0.1183	0.0013	
660	0.2883	0.0119		0.0454	0.0006	
700	0.2813	0.0117		0.1559	0.0007	

CALIBRATION PLANNING

Test name	Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
+THC 50	DAT Cal	*								
+Amphet 500	DAT Cal	*								
+Benzodiazep300	DAT Cal	*								
+Cocaine 300	DAT Cal	*								
+Opiate 300	DAT Cal	*								

MAINTENANCE PLANNING

Maintenance procedure	0	1	2	3	4	5	6	7	>7 days valid
NEEDLE RINSE								*	
SYSTEM CLEAN									*

EXPIRING REAGENTS

Reagent name	0	1	2	3	4	5	6	7	>7 days valid

EXPIRING CONTROLS

Control name	0	1	2	3	4	5	6	7	>7 days valid
Level 0									*
Level 5						*			
S2E									*
S2S									*

EXPIRING CALIBRATORS

Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
Acetaminophen									*
Salicylate Cal									*
sTOX Cal									*
sTCA Cal									*
ETOH Cal									*
DAT Cal									*
Ecstasy Cal 2									*
Ecstasy Cal 3									*
LSD Cal									*
qEcstasyCal 300									*
qEcstasyCal 500									*
qLSD Cal									*
cR Cal									*
sG Cal									*
pH Cal									*
cH Cal									*
nT Cal									*
oX Cal									*

Measurement date: 1/25/2017 11:46:30 AM

Control name: Level 0

Batch number: 9a508ul-j3

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	negative	0.338 dAbs/m		0.050 dAbs/m	-
+Benzodiazep300	negative	0.287 dAbs/m		0.112 dAbs/m	-
+THC 50	negative	0.287 dAbs/m		0.044 dAbs/m	-
+Cocaine 300	negative	0.236 dAbs/m		0.110 dAbs/m	-
+Opiate 300	negative	0.191 dAbs/m		0.078 dAbs/m	-

Measurement date: 1/25/2017 11:58:12 AM

Control name: Level 5

Batch number: 9A609UL-h3

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.542 dAbs/m		0.154 dAbs/m	+
+Benzodiazep300	POSITIVE	0.513 dAbs/m		0.114 dAbs/m	+
+THC 50	POSITIVE	0.478 dAbs/m		0.147 dAbs/m	+
+Cocaine 300	POSITIVE	0.416 dAbs/m		0.071 dAbs/m	+
+Opiate 300	POSITIVE	0.428 dAbs/m		0.159 dAbs/m	+

Measurement date: 1/25/2017 11:59:59 AM

Control name: S2E

Batch number: 72520

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Benzodiazep300	POSITIVE	0.453 dAbs/m		0.053 dAbs/m	+
+THC 50	POSITIVE	0.418 dAbs/m		0.087 dAbs/m	+
+Cocaine 300	POSITIVE	0.389 dAbs/m		0.043 dAbs/m	+
+Opiate 300	POSITIVE	0.353 dAbs/m		0.084 dAbs/m	+

Measurement date: 1/25/2017 12:00:52 PM

Control name: S2S

Batch number: 60100

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.516 dAbs/m		0.128 dAbs/m	+



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen
Analytical Method 1.1

Laboratory Case No: neg-121616
Date of Analysis: 1/25/2017
Time of Analysis: 12:13:30 PM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	negative 0.336	0.388
Benzodiazepines (300 ng/mL)	negative 0.288	0.400
Cannabinoids (50 ng/mL)	negative 0.283	0.331
Cocaine Metabolite (300 ng/mL)	negative 0.235	0.345
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	negative 0.190	0.269

VIVA-Jr Instrument ID: 8-1324



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen
Analytical Method 1.1

Laboratory Case No: brc3-68430
Date of Analysis: 1/25/2017
Time of Analysis: 12:01:48 PM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.410	0.388
Benzodiazepines (300 ng/mL)	POSITIVE 0.638	0.400
Cannabinoids (50 ng/mL)	negative 0.304	0.331
Cocaine Metabolite (300 ng/mL)	negative 0.293	0.345
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.456	0.269

VIVA-Jr Instrument ID: 8-1324



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen
Analytical Method 1.1

Laboratory Case No: brs3-67140
Date of Analysis: 1/25/2017
Time of Analysis: 12:04:03 PM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.506	0.388
Benzodiazepines (300 ng/mL)	POSITIVE 0.644	0.400
Cannabinoids (50 ng/mL)	POSITIVE 0.425	0.331
Cocaine Metabolite (300 ng/mL)	POSITIVE 0.370	0.345
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.355	0.269

VIVA-Jr Instrument ID: 8-1324