






















*BWylee*

**Worklist: 1678**

<u>LAB CASE</u>	<u>ITEM</u>	<u>TASK ID</u>	<u>DESCRIPTION</u>	
C2017-0659	2	81180	AM 1 Urine Screening EMIT	
C2017-0663	1	81270	AM 1 Urine Screening EMIT	
C2017-0679	1	81572	AM 1 Urine Screening EMIT	
C2017-0684	1	81606	AM 1 Urine Screening EMIT	
C2017-0687	1	81684	AM 1 Urine Screening EMIT	
C2017-0699	1	81735	AM 1 Urine Screening EMIT	
C2017-0752	1	82377	AM 1 Urine Screening EMIT	
C2017-0753	1	82380	AM 1 Urine Screening EMIT	
C2017-0766	1	82425	AM 1 Urine Screening EMIT	
C2017-0776	1	82669	AM 1 Urine Screening EMIT	
C2017-0781	1	82895	AM 1 Urine Screening EMIT	
M2017-1246	10	82771	AM 1 Urine Screening EMIT	
M2017-1246	13	82776	AM 1 Urine Screening EMIT	
M2017-1246	2	82758	AM 1 Urine Screening EMIT	
M2017-1246	4	82761	AM 1 Urine Screening EMIT	
M2017-1246	6	82765	AM 1 Urine Screening EMIT	
M2017-1246	8	82768	AM 1 Urine Screening EMIT	
M2017-1327	1	82822	AM 1 Urine Screening EMIT	
M2017-1681	3	82389	AM 1 Urine Screening EMIT	
P2017-0248	4	81161	AM 1 Urine Screening EMIT	
P2017-0248	5	81165	AM 1 Urine Screening EMIT	

*[Handwritten signature]*

### AUTO MAINTENANCE

### BLANK ROTOR RESULTS

Blank date/time: 4/25/2017 4:15:29 AM

Wavelength	Cuvette AV	Cuvette SD	Errors	Lamp AV	Lamp SD	Errors
340	0.3931	0.0164		1.9381	0.0009	
415	0.3508	0.0062		0.7867	0.0005	
505	0.2576	0.0058		0.2184	0.0003	
546	0.2402	0.0056		0.1476	0.0005	
570	0.2448	0.0057		0.1584	0.0004	
600	0.2663	0.0056		0.0334	0.0011	
660	0.2653	0.0055		-0.0080	0.0006	
700	0.2905	0.0054		0.0611	0.0006	

### CALIBRATION PLANNING

Test name	Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
+THC 50	DAT Cal	*								
+Amphet 500	DAT Cal	*								
+Benzodiazep300	DAT Cal	*								
+Cocaine 300	DAT Cal	*								
+Opiate 300	DAT Cal	*								

### MAINTENANCE PLANNING

Maintenance procedure	0	1	2	3	4	5	6	7	>7 days valid
NEEDLE RINSE								*	
SYSTEM CLEAN									*

### EXPIRING REAGENTS

Reagent name	0	1	2	3	4	5	6	7	>7 days valid

### EXPIRING CONTROLS

Control name	0	1	2	3	4	5	6	7	>7 days valid
Level 0									*
Level 5									*
S2E									*
S2S									*

### EXPIRING CALIBRATORS

Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
Acetaminophen									*
Salicylate Cal									*
sTOX Cal									*
sTCA Cal									*
ETOH Cal									*
DAT Cal									*
Ecstasy Cal 2									*
Ecstasy Cal 3									*
LSD Cal									*
qEcstasyCal 300									*
qEcstasyCal 500									*
qLSD Cal									*
cR Cal									*
sG Cal									*
pH Cal									*
ch Cal									*
nT Cal									*
oX Cal									*

Measurement date: 4/25/2017 8:55:45 AM

Control name: Level 0

Batch number: 9a508ul-j3

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	negative	0.343 dAbs/m	0.037	dAbs/m	-
+Benzodiazep300	negative	0.283 dAbs/m	0.118	dAbs/m	-
+THC 50	negative	0.286 dAbs/m	0.037	dAbs/m	-
+Cocaine 300	negative	0.249 dAbs/m	0.114	dAbs/m	-
+Opiate 300	negative	0.190 dAbs/m	0.079	dAbs/m	-

Measurement date: 4/25/2017 8:58:01 AM

Control name: Level 5

Batch number: 9A609UL-J4

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.540 dAbs/m	0.161	dAbs/m	+
+Benzodiazep300	POSITIVE	0.493 dAbs/m	0.092	dAbs/m	+
+THC 50	POSITIVE	0.473 dAbs/m	0.150	dAbs/m	+
+Cocaine 300	POSITIVE	0.427 dAbs/m	0.064	dAbs/m	+
+Opiate 300	POSITIVE	0.419 dAbs/m	0.150	dAbs/m	+

Measurement date: 4/25/2017 8:59:48 AM

Control name: S2E

Batch number: 72520

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Benzodiazep300	POSITIVE	0.427 dAbs/m	0.025	dAbs/m	+
+THC 50	POSITIVE	0.347 dAbs/m	0.024	dAbs/m	+
+Cocaine 300	POSITIVE	0.390 dAbs/m	0.027	dAbs/m	+
+Opiate 300	POSITIVE	0.334 dAbs/m	0.066	dAbs/m	+

Measurement date: 4/25/2017 9:08:46 AM

Control name: S2S

Batch number: 60100

Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.478 dAbs/m	0.099	dAbs/m	+



## Idaho State Police Forensic Services

**EMIT Urine Drugs-of-Abuse Screen**  
Analytical Method 1.1

Laboratory Case No: neg 31517  
Date of Analysis: 4/25/2017  
Time of Analysis: 11:06:06 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	negative 0.346	0.380
Benzodiazepines (300 ng/mL)	negative 0.278	0.401
Cannabinoids (50 ng/mL)	negative 0.287	0.324
Cocaine Metabolite (300 ng/mL)	negative 0.257	0.363
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	negative 0.191	0.269

VIVA-Jr Instrument ID: 8-1324



## Idaho State Police Forensic Services

### EMIT Urine Drugs-of-Abuse Screen

Analytical Method 1.1

Laboratory Case No: <sup>68430</sup> brc3-~~67140~~ <sup>A</sup>  
 Date of Analysis: 4/25/2017  
 Time of Analysis: 10:56:38 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.411	0.380
Benzodiazepines (300 ng/mL)	POSITIVE 0.642	0.401
Cannabinoids (50 ng/mL)	negative 0.308	0.324
Cocaine Metabolite (300 ng/mL)	negative 0.294	0.363
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.449	0.269

VIVA-Jr Instrument ID: 8-1324



## Idaho State Police Forensic Services

### EMIT Urine Drugs-of-Abuse Screen Analytical Method 1.1

Laboratory Case No: ~~brs 68430~~ <sup>CA</sup> <sup>A</sup> brs3 ~~47~~ 67140  
 Date of Analysis: 4/25/2017  
 Time of Analysis: 10:54:24 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.509	0.380
Benzodiazepines (300 ng/mL)	POSITIVE 0.621	0.401
Cannabinoids (50 ng/mL)	POSITIVE 0.412	0.324
Cocaine Metabolite (300 ng/mL)	POSITIVE 0.387	0.363
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.356	0.269

VIVA-Jr Instrument ID: 8-1324