













Bylee

Worklist: 1702

<u>LAB CASE</u>	<u>ITEM</u>	<u>TASK ID</u>	<u>DESCRIPTION</u>	
C2017-0747	1	82333	AM 1 Urine Screening EMIT	
C2017-0783	1	82985	AM 1 Urine Screening EMIT	
C2017-0784	1	82988	AM 1 Urine Screening EMIT	
C2017-0793	1	83097	AM 1 Urine Screening EMIT	
C2017-0796	1	83127	AM 1 Urine Screening EMIT	
C2017-0800	4	83268	AM 1 Urine Screening EMIT	
C2017-0802	1	83283	AM 1 Urine Screening EMIT	
C2017-0803	1	83286	AM 1 Urine Screening EMIT	
C2017-0811	1	83464	AM 1 Urine Screening EMIT	
C2017-0812	1	83467	AM 1 Urine Screening EMIT	
C2017-0813	1	83470	AM 1 Urine Screening EMIT	
M2017-1797	3	83090	AM 1 Urine Screening EMIT	



AUTO MAINTENANCE**BLANK ROTOR RESULTS**

Blank date/time: 5/1/2017 9:26:15 AM

Wavelength	Cuvette AV	Cuvette SD	Errors	Lamp AV	Lamp SD	Errors
340	0.3936	0.0172		1.9396	0.0006	
415	0.3152	0.0064		0.7864	0.0007	
505	0.2578	0.0057		0.2191	0.0003	
546	0.2407	0.0056		0.1486	0.0004	
570	0.2455	0.0055		0.1593	0.0004	
600	0.2667	0.0055		0.0356	0.0014	
660	0.2656	0.0054		-0.0056	0.0007	
700	0.2853	0.0052		0.0618	0.0006	

CALIBRATION PLANNING

Test name	Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
+THC 50	DAT Cal	*								
+Amphet 500	DAT Cal	*								
+Benzodiazep300	DAT Cal	*								
+Cocaine 300	DAT Cal	*								
+Opiate 300	DAT Cal	*								

MAINTENANCE PLANNING

Maintenance procedure	0	1	2	3	4	5	6	7	>7 days valid
NEEDLE RINSE								*	
SYSTEM CLEAN									*

EXPIRING REAGENTS

Reagent name	0	1	2	3	4	5	6	7	>7 days valid

EXPIRING CONTROLS

Control name	0	1	2	3	4	5	6	7	>7 days valid
Level 0									*
Level 5									*
S2E									*
S2S									*

EXPIRING CALIBRATORS

Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
Acetaminophen									*
Salicylate Cal									*
sTOX Cal									*
sTCA Cal									*
ETOH Cal									*
DAT Cal									*
Ecstasy Cal 2									*
Ecstasy Cal 3									*
LSD Cal									*
qEcstasyCal 300									*
qEcstasyCal 500									*
qLSD Cal									*
cR Cal									*
sG Cal									*
pH Cal									*
ch Cal									*
nT Cal									*
oX Cal									*

Measurement date: 5/1/2017 9:45:21 AM
 Control name: Level 0
 Batch number: 9a508ul-j3
 Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	negative	0.339 dAbs/m	0.041	dAbs/m	-
+Benzodiazep300	negative	0.282 dAbs/m	0.119	dAbs/m	-
+THC 50	negative	0.289 dAbs/m	0.035	dAbs/m	-
+Cocaine 300	negative	0.247 dAbs/m	0.116	dAbs/m	-
+Opiate 300	negative	0.192 dAbs/m	0.077	dAbs/m	-

Measurement date: 5/1/2017 9:47:36 AM
 Control name: Level 5
 Batch number: 9A609UL-J4
 Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.549 dAbs/m	0.170	dAbs/m	+
+Benzodiazep300	POSITIVE	0.500 dAbs/m	0.099	dAbs/m	+
+THC 50	POSITIVE	0.483 dAbs/m	0.160	dAbs/m	+
+Cocaine 300	POSITIVE	0.425 dAbs/m	0.062	dAbs/m	+
+Opiate 300	POSITIVE	0.434 dAbs/m	0.165	dAbs/m	+

Measurement date: 5/1/2017 9:49:24 AM
 Control name: S2E
 Batch number: 72520
 Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Benzodiazep300	POSITIVE	0.449 dAbs/m	0.048	dAbs/m	+
+THC 50	POSITIVE	0.376 dAbs/m	0.053	dAbs/m	+
+Cocaine 300	POSITIVE	0.395 dAbs/m	0.032	dAbs/m	+
+Opiate 300	POSITIVE	0.343 dAbs/m	0.075	dAbs/m	+

Measurement date: 5/1/2017 9:58:22 AM
 Control name: S2S
 Batch number: 60100
 Sample type: Control READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.491 dAbs/m	0.112	dAbs/m	+





Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen
Analytical Method 1.1

Laboratory Case No: brs3 67140
Date of Analysis: 5/1/2017
Time of Analysis: 10:53:41 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.512	0.380
Benzodiazepines (300 ng/mL)	POSITIVE 0.626	0.401
Cannabinoids (50 ng/mL)	POSITIVE 0.404	0.324
Cocaine Metabolite (300 ng/mL)	POSITIVE 0.387	0.363
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.362	0.269

VIVA-Jr Instrument ID: 8-1324



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen
Analytical Method 1.1

Laboratory Case No: brc3 68430
Date of Analysis: 5/1/2017
Time of Analysis: 10:55:56 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	POSITIVE 0.416	0.380
Benzodiazepines (300 ng/mL)	POSITIVE 0.619	0.401
Cannabinoids (50 ng/mL)	negative 0.298	0.324
Cocaine Metabolite (300 ng/mL)	negative 0.304	0.363
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.457	0.269

VIVA-Jr Instrument ID: 8-1324



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen
Analytical Method 1.1

Laboratory Case No: neg 42617
Date of Analysis: 5/1/2017
Time of Analysis: 11:05:24 AM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	negative 0.332	0.380
Benzodiazepines (300 ng/mL)	negative 0.279	0.401
Cannabinoids (50 ng/mL)	negative 0.281	0.324
Cocaine Metabolite (300 ng/mL)	negative 0.250	0.363
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	negative 0.193	0.269

VIVA-Jr Instrument ID: 8-1324