




















9/11/2017

Worklist: 1886

<u>LAB CASE</u>	<u>ITEM</u>	<u>TASK ID</u>	<u>DESCRIPTION</u>	
C2017-1447	1	94533	AM 1 Urine Screening EMIT	
C2017-1464	2	94519	AM 1 Urine Screening EMIT	
C2017-1483	2	94493	AM 1 Urine Screening EMIT	
C2017-1591	1	91893	AM 1 Urine Screening EMIT	
C2017-1598	1	94502	AM 1 Urine Screening EMIT	
C2017-1684	2	92834	AM 1 Urine Screening EMIT	
C2017-1733	2	94527	AM 1 Urine Screening EMIT	
C2017-1747	3	94534	AM 1 Urine Screening EMIT	
C2017-1748	2	94518	AM 1 Urine Screening EMIT	
C2017-1763	1	94525	AM 1 Urine Screening EMIT	
C2017-1768	1	94531	AM 1 Urine Screening EMIT	
C2017-1781	1	94512	AM 1 Urine Screening EMIT	
C2017-1797	1	94513	AM 1 Urine Screening EMIT	
C2017-1805	1	94495	AM 1 Urine Screening EMIT	
C2017-1806	1	94494	AM 1 Urine Screening EMIT	
M2017-3374	2	94492	AM 1 Urine Screening EMIT	
M2017-3409	3	94491	AM 1 Urine Screening EMIT	

BW

9/8/17 12:00 PM
aw

AUTO MAINTENANCE

BLANK ROTOR RESULTS

Blank date/time: 9/6/2017 1:08:40 AM

Wavelength	Cuvette AV	Cuvette SD	Errors	Lamp AV	Lamp SD	Errors
340	0.4033	0.0066		1.9503	0.0011	
415	0.3520	0.0056		0.7869	0.0007	
505	0.2591	0.0053		0.2217	0.0004	
546	0.2416	0.0052		0.1505	0.0005	
570	0.2464	0.0052		0.1616	0.0005	
600	0.2678	0.0051		0.0349	0.0018	
660	0.2669	0.0050		-0.0053	0.0006	
700	0.2920	0.0049		0.0628	0.0006	

CALIBRATION PLANNING

Test name	Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
+THC 50	DAT Cal	*								
+Amphet 500	DAT Cal	*								
+Benzodiazep300	DAT Cal	*								
+Cocaine 300	DAT Cal	*								
+Opiate 300	DAT Cal	*								

MAINTENANCE PLANNING

Maintenance procedure	0	1	2	3	4	5	6	7	>7 days valid
NEEDLE RINSE	*								
SYSTEM CLEAN	*								

EXPIRING REAGENTS

Reagent name	0	1	2	3	4	5	6	7	>7 days valid

EXPIRING CONTROLS

Control name	0	1	2	3	4	5	6	7	>7 days valid
Level 0	*								
Level 5	*								
S2E									*
S2S									*

* > updated after blank
BW

EXPIRING CALIBRATORS

Calibrator name	0	1	2	3	4	5	6	7	>7 days valid
Acetaminophen									*
Salicylate Cal									*
sTOX Cal									*
sTCA Cal									*
ETOH Cal									*
DAT Cal									*
Ecstasy Cal 2									*
Ecstasy Cal 3									*
LSD Cal									*
qEcstasyCal 300									*
qEcstasyCal 500									*
qLSD Cal									*
cR Cal									*
sG Cal									*
pH Cal									*
cH Cal									*
nT Cal									*
oX Cal									*

BW

BLANK/CALIBRATOR INFO

9/6/2017
6:00
12:40 pm

Test name	Reagent blank	Calibrator	
+Amphet 500		Accepted	Yes
		Value	0.4501 dAbs/m
		Meas. date	9/6/2017
		Type	One point - Cut-off
+Benzodiazep300		Accepted	Yes
		Value	0.4131 dAbs/m
		Meas. date	9/6/2017
		Type	One point - Cut-off
+THC 50		Accepted	Yes
		Value	0.3475 dAbs/m
		Meas. date	9/6/2017
		Type	One point - Cut-off
+Cocaine 300		Accepted	Yes
		Value	0.3708 dAbs/m
		Meas. date	9/6/2017
		Type	One point - Cut-off
+Opiate 300		Accepted	Yes
		Value	0.3109 dAbs/m
		Meas. date	9/6/2017
		Type	One point - Cut-off

BW

Measurement date: 9/6/2017 1:50:19 AM
 Control name: Level 0
 Batch number: 9a508ul-k1
 Sample type: Control

READY

9/8/17 12:50 PM
 BL

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	negative	0.331 dAbs/m		0.119 dAbs/m	-
+Benzodiazep300	negative	0.285 dAbs/m		0.128 dAbs/m	-
+THC 50	negative	0.282 dAbs/m		0.065 dAbs/m	-
+Cocaine 300	negative	0.243 dAbs/m		0.128 dAbs/m	-
+Opiate 300	negative	0.189 dAbs/m		0.122 dAbs/m	-

Measurement date: 9/6/2017 1:52:35 AM
 Control name: Level 5
 Batch number: 9A609UL-k1
 Sample type: Control

READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.529 dAbs/m		0.079 dAbs/m	+
+Benzodiazep300	POSITIVE	0.490 dAbs/m		0.076 dAbs/m	+
+THC 50	POSITIVE	0.476 dAbs/m		0.128 dAbs/m	+
+Cocaine 300	POSITIVE	0.403 dAbs/m		0.032 dAbs/m	+
+Opiate 300	POSITIVE	0.426 dAbs/m		0.115 dAbs/m	+

Measurement date: 9/6/2017 1:54:23 AM
 Control name: S2E
 Batch number: 72520
 Sample type: Control

READY

Test name	Result	Units	Separation	Units	Flags
+Benzodiazep300	POSITIVE	0.435 dAbs/m		0.022 dAbs/m	+
+THC 50	POSITIVE	0.370 dAbs/m		0.023 dAbs/m	+
+Cocaine 300	POSITIVE	0.381 dAbs/m		0.010 dAbs/m	+
+Opiate 300	POSITIVE	0.339 dAbs/m		0.028 dAbs/m	+

Measurement date: 9/6/2017 1:56:09 AM
 Control name: S2S
 Batch number: 60130
 Sample type: Control

READY

Test name	Result	Units	Separation	Units	Flags
+Amphet 500	POSITIVE	0.461 dAbs/m		0.010 dAbs/m	+

WORKLIST

Pos	Type	Rep	Sample data	Tests requested		
1	Sample	1	ID: brc3-68440 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
2	Sample	1	ID: brs2s 60130 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
3	Sample	1	ID: neg 9517 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
4	Sample	1	ID: c2017-1447-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
5	Sample	1	ID: c2017-1464-2 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
6	Sample	1	ID: c2017-1483-2 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
7	Sample	1	ID: c2017-1529-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
8	Sample	1	ID: c2017-1591-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
9	Sample	1	ID: c2017-1598-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
10	Sample	1	ID: c2017-1684-2 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
11	Sample	1	ID: c2017-1733-2 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
12	Sample	1	ID: c2017-1747-3 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
13	Sample	1	ID: c2017-1748-2 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
14	Sample	1	ID: c2017-1763-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
15	Sample	1	ID: c2017-1768-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
16	Sample	1	ID: c2017-1781-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
17	Sample	1	ID: c2017-1797-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
18	Sample	1	ID: c2017-1805-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
19	Sample	1	ID: c2017-1806-1 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
20	Sample	1	ID: m2017-3374-2 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
21	Sample	1	ID: m2017-3409-3 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
22	Sample	1	ID: m2017-3448-2 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
23	Sample	1	ID: m2017-3532-2 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
24	Sample	1	ID: m2017-3538-3 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50
25	Sample	1	ID: m2017-3791-3 Name:	1 +Amphet 500 4 +Cocaine 300	2 +Benzodiazep300 5 +Opiate 300	3 +THC 50



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen

Analytical Method 1.1

Laboratory Case No: neg 9517
Date of Analysis: 9/8/2017
Time of Analysis: 3:10:49 PM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	negative 0.321	0.450
Benzodiazepines (300 ng/mL)	negative 0.275	0.413
Cannabinoids (50 ng/mL)	negative 0.275	0.348
Cocaine Metabolite (300 ng/mL)	negative 0.243	0.371
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	negative 0.191	0.311

VIVA-Jr Instrument ID: 8-1324



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen
Analytical Method 1.1

Laboratory Case No: brc3-68440
Date of Analysis: 9/8/2017
Time of Analysis: 2:45:40 PM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	negative 0.430 <i>elevated response OK BW</i>	0.450
Benzodiazepines (300 ng/mL)	POSITIVE 0.619	0.413
Cannabinoids (50 ng/mL)	negative 0.297	0.348
Cocaine Metabolite (300 ng/mL)	negative 0.284	0.371
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.450	0.311

VIVA-Jr Instrument ID: 8-1324



Idaho State Police Forensic Services

EMIT Urine Drugs-of-Abuse Screen

Analytical Method 1.1

Laboratory Case No: brs2s 60130

Date of Analysis: 9/8/2017

Time of Analysis: 2:47:55 PM

Assay	Results/Interpretation	Cutoff Value (absorbance)
Amphetamines (500 ng/mL)	negative 0.432	0.450
Benzodiazepines (300 ng/mL)	negative 0.280	0.413
Cannabinoids (50 ng/mL)	POSITIVE 0.355	0.348
Cocaine Metabolite (300 ng/mL)	negative 0.323	0.371
Methadone (300 ng/mL)		
Opiates (300 ng/mL)	POSITIVE 0.391	0.311

VIVA-Jr Instrument ID: 8-1324