



Idaho State Police Forensic Services

CERTIFICATE OF ANALYSIS/APPROVAL

The Idaho State Police Forensic Services (ISPFS) hereby certifies and approves Alcohol Simulator Solution **Lot Number 19803** (a product manufactured by REPCO Marketing, Inc.) to be used to conduct performance verification checks within the State of Idaho in accordance with the analytical methods, policies and/or procedures promulgated by the Department governing breath alcohol examinations. ISPFS also approves of the manufacturer of this solution (REPCO Marketing, Inc.) to provide Alcohol Simulator Solution **Lot Number 19803** in the State of Idaho. This lot has a target value of 0.080 with a range of 0.072 to 0.088 grams of ethyl alcohol/210 liters of vapor.

The expiration date for this lot number is on July 10th, 2021 at 11:59 PM.

8-14-19

Date

Volatiles Analysis Discipline Leader

STATE OF IDAHO)
) ss.
County of Kootenai)

On this 14th day of August, in the year 2019, before me, Lynn A. Higdem, a notary public, Jeremy Johnston personally appeared, known to me to be the person whose name is subscribed to the within instrument as a Forensic Scientist for the Idaho State Police Forensic Services, and acknowledged to me that he executed the same as such Scientist.

Lynn A. Higdem Notary Public

My Commission Expires: Sept. 8, 2022

