



Idaho State Police Forensic Services

CERTIFICATE OF ANALYSIS/APPROVAL

The Idaho State Police Forensic Services (ISPFS) hereby certifies and approves Alcohol Simulator Solution **Lot Number 18110** (a product manufactured by RepCo Marketing, Inc.) to be used to conduct performance verification checks within the State of Idaho in accordance with the analytical methods, policies and/or procedures promulgated by the Department governing breath alcohol examinations. ISPFS also approves of the manufacturer of this solution (RepCo Marketing, Inc.) to provide Alcohol Simulator Solution **Lot Number 18110** in the State of Idaho. This lot has a target value of 0.200 with a range of 0.180 to 0.220 grams of ethyl alcohol/210 liters of vapor.

The expiration date for this lot number is on October 16th, 2020 at 11:59 PM.

12-3-18

Date

Volatiles Analysis Discipline Leader

STATE OF IDAHO)
) ss.
County of Kootenai)

On this 3rd day of December, in the year 2018, before me, Dorie Mallory, a notary public, Jeremy Johnston personally appeared, known to me to be the person whose name is subscribed to the within instrument as a Forensic Scientist for the Idaho State Police Forensic Services, and acknowledged to me that he executed the same as such Scientist.

Dorie Mallory Notary Public

My Commission Expires: 5/13/20

