



Idaho State Police Forensic Services

CERTIFICATE OF ANALYSIS

The Idaho State Police Forensic Services hereby certifies that Alcohol Simulator Solution Lot Number 09806 to be used to conduct calibration checks within the State of Idaho in accordance with the policies and/or procedures promulgated by the Department governing breath alcohol examinations. This lot has a target value of 0.080 with a range of 0.072 to 0.088 grams of ethyl alcohol/210 liters of vapor.

The expiration date for this lot number is November 25, 2011 at 11:59 PM.

March 3, 2010
Date


Breath Alcohol Discipline Leader

STATE OF IDAHO)
) ss.
County of Ada)

On this 3rd day of March, in the year 2010, before me, Judy Packer, a notary public, personally appeared Darren B. Jewkes, known to me to be the person whose name is subscribed to the within instrument as a Forensic Scientist for the Idaho State Police Forensic Services, and acknowledge to me that he executed the same as such Scientist.


Notary Public

My Commission Expires: 4 October 2013

