




*Idaho State Police Forensic Services*

**CERTIFICATE OF APPROVAL**

The Idaho State Police Forensic Services (ISPFS) hereby certifies and approves Alcohol Simulator Solution **Lot Number 14803** (a product manufactured by RepCo Marketing Inc.) to be used to conduct performance verification checks within the State of Idaho in accordance with the analytical methods, policies and/or procedures promulgated by the Department governing breath alcohol examinations. ISPFS also approves of the manufacturer of this solution (RepCo Marketing Inc.) to provide Alcohol Simulator Solution **Lot Number 14803** in the State of Idaho. This lot has a target value of 0.080 with a range of 0.072 to 0.088 grams of ethyl alcohol/210 liters of vapor.

The expiration date for this lot number is on September 10<sup>th</sup>, 2016 at 11:59 PM.

11-3-14  
Date

  
Volatiles Analysis Discipline Leader

STATE OF IDAHO )  
                          ) ss.  
County of Kootenai)

On this 3rd day of November, in the year 2014, before me, Lindsay Stores, a notary public, Jeremy Johnston personally appeared, known to me to be the person whose name is subscribed to the within instrument as a Forensic Scientist for the Idaho State Police Forensic Services, and acknowledged to me that he executed the same as such Scientist.

 Notary Public  
My Commission Expires: 1.22.2019

